

Flu A/B/RSV Assay (Panther Fusion™ System)

For *in vitro* diagnostic use.

For U.S. Export Only.

CONTENTS

- General Information 2**
 - Intended Use 2
 - Summary and Explanation of the Test 2
 - Principles of the Procedure 3
 - Warnings and Precautions 4
 - Reagent Storage and Handling Requirements 6
 - Specimen Collection and Storage 7
 - Specimen Transport 8
- Panther Fusion System 9**
 - Reagents and Materials Provided for the Panther Fusion Flu A/B/RSV Assay 9
 - Materials Required and Available Separately 10
 - Panther Fusion System Test Procedure 11
 - Procedural Notes 12
- Quality Control 12**
- Interpretation of Results 13**
- Limitations 14**
- Panther Fusion System Assay Performance 15**
 - Clinical Performance 15
 - Analytical Sensitivity 16
 - Reactivity 16
 - Analytical Specificity 18
 - Competitive Interference 20
 - Interference 21
 - Carry-Over/Contamination 21
 - Assay Precision 22
- Bibliography 24**

General Information

Intended Use

The Panther Fusion™ Flu A/B/RSV assay is a multiplex real-time PCR (RT-PCR) *in vitro* diagnostic test for the rapid and qualitative detection and differentiation of influenza A virus, influenza B virus, and respiratory syncytial virus (RSV). Nucleic acids are isolated and purified from nasopharyngeal (NP) swab specimens obtained from individuals exhibiting signs and symptoms of a respiratory tract infection.

This assay is intended to aid in the differential diagnosis of influenza A virus, influenza B virus and RSV infections in humans and is not intended to detect influenza C virus infections. Negative results do not preclude influenza A virus, influenza B virus or RSV infections and should not be used as the sole basis for treatment or other management decisions. This assay is designed for use on the Panther Fusion system.

Summary and Explanation of the Test

Respiratory viruses are responsible for a wide range of acute respiratory tract infections including the common cold, influenza, and croup and represent the most common cause of acute illness in the United States. Disease severity can be especially high in the young, the immunocompromised, and elderly patients. Accurate and timely diagnosis of the cause of respiratory tract infections has many benefits. They include improved treatment of the patient by ensuring appropriate antiviral treatment (e.g. oseltamivir for influenza), decreasing the overall cost of care, reducing selection for antimicrobial resistant organisms due to excessive and inappropriate use of antibiotics,¹ assisting infection control personnel in providing appropriate measures to minimize nosocomial spread, and providing valued information to public health authorities regarding which viruses are circulating in the community.²

Influenza is an acute respiratory illness caused by infection with the influenza virus, primarily types A and B.³ Influenza A viruses are further categorized into subtypes based on the two major surface protein antigens: hemagglutinin (H) and neuraminidase (N).⁴ Influenza B viruses are not categorized into subtypes.⁴ Influenza viruses continuously undergo genetic changes including drift (random mutation) and variation (genomic reassortment), generating new strains of virus each year, leaving the human population vulnerable to these seasonal changes. Epidemics occur yearly (typically in winter) and while both types A and B circulate in the population, type A is usually dominant. Transmission of influenza is primarily via airborne droplet (coughing or sneezing). Symptoms arise on average 1 to 2 days post-exposure and include fever, chills, headache, malaise, cough, and coryza.

Complications due to influenza include pneumonia causing increased morbidity and mortality in pediatric, elderly and immunocompromised populations. Influenza occurs globally with an annual attack rate estimated at 5%–10% in adults and 20%–30% in children. Illnesses can result in hospitalization and death mainly among high-risk groups (the very young, elderly or chronically ill). Worldwide, these annual epidemics are estimated to result in about 3 to 5 million cases of severe illness, and about 250,000 to 500,000 deaths.⁵

Respiratory syncytial virus (RSV) is a leading cause of respiratory infections in infants and children. There are 2 types of RSV (A and B) based on antigenic and surface protein variations.

Most yearly epidemics (typically during winter) contain a mix of type A and B viruses, but one subgroup can dominate during a season. RSV infection can cause severe respiratory illness among all ages but is more prevalent in pediatric, elderly and immunocompromised populations. Annually in the United States, RSV infection has been associated with an estimated 57,527 hospitalizations and 2.1 million outpatient visits among children younger than 5 years, and 177,000 hospitalizations and 14,000 deaths among adults older than 65 years.⁶

Principles of the Procedure

The Panther Fusion Flu A/B/RSV assay involves the following steps: sample lysis, nucleic acid capture and elution transfer, and multiplex RT-PCR when analytes are simultaneously amplified, detected and differentiated. Nucleic acid capture and elution takes place in a single tube on the Panther Fusion system. The eluate is transferred to the Panther Fusion system reaction tube containing the assay reagents. Multiplex RT-PCR is then performed for the eluted nucleic acid on the Panther Fusion system.

Nucleic acid capture and elution: Prior to processing and testing on the Panther Fusion system, specimens are transferred to a Specimen Lysis Tube containing specimen transport media (STM) that lyses the cells, releases target nucleic acid and protects them from degradation during storage.

The Internal Control-S (IC-S) is added to each test specimen and controls via the working Panther Fusion Capture Reagent-S (wFCR-S). The IC-S in the reagent monitors specimen processing, amplification and detection.

Capture oligonucleotides hybridize to nucleic acid in the test specimen. Hybridized nucleic acid is then separated from the specimen in a magnetic field.

Wash steps remove extraneous components from the reaction tube. The elution step elutes purified nucleic acid. During the nucleic acid capture and elution step, total nucleic acid is isolated from specimens.

Elution transfer and RT-PCR: During the elution transfer step, eluted nucleic acid is transferred to a Panther Fusion reaction tube already containing oil and reconstituted mastermix.

Target amplification occurs via RT-PCR. A reverse transcriptase generates a DNA copy of the target sequence. Target specific forward and reverse primers and probes then amplify targets while simultaneously detecting and discriminating multiple target types via multiplex RT-PCR.

The Panther Fusion system compares the fluorescence signal to a predetermined cut-off to produce a qualitative result for the presence or absence of the analyte.

The analytes and the channel used for their detection on the Panther Fusion system is summarized in the table below.

Analyte	Gene Targeted	Instrument Channel
Influenza A Virus	Matrix	FAM
Respiratory Syncytial Virus A/B	Matrix	HEX
Influenza B Virus	Matrix	ROX
Internal Control	Not applicable	RED677




Warnings and Precautions

- A. For *in vitro* diagnostic use.
- B. Carefully read this entire package insert and the *Panther Fusion System Operator's Manual*.
- C. The Panther Fusion Enhancer Reagent-S (FER-S) is corrosive, harmful if swallowed and causes severe skin burns and eye damage.
- D. Only personnel adequately trained on the use of this assay and in handling potentially infectious materials should perform these procedures. If a spill occurs, immediately disinfect using appropriate site procedures.
- E. Handle all specimens as if infectious, using safe laboratory procedures such as those outlined in CDC/NIH Biosafety in Microbiological and Biomedical Laboratories and in the CLSI Document M29 Protection of Laboratory Workers from Occupationally Acquired Infections.

Note: *If infection with a novel influenza A virus is suspected based on current clinical and epidemiological screening criteria recommended by public health authorities, collect specimens with appropriate infection control precautions for novel virulent influenza viruses and send to state or local health department for testing. Do not attempt viral culture in these cases unless a BSL 3+ facility is available to receive and culture specimens.*

- F. Use only supplied or specified disposable laboratory ware.
- G. Wear disposable, powderless gloves, protective eye wear, and laboratory coats when handling specimens and reagents. Wash hands thoroughly after handling specimens and reagents.
- H. Dispose of all material that has come into contact with specimens and reagents in accordance with applicable national, international, and regional regulations.
- I. Expiration dates listed on the Panther Fusion Specimen Lysis Tubes pertains to the transfer of sample into the tube and not to testing of the sample. Specimens collected/transferred any time prior to these expiration dates are valid for testing provided they are transported and stored in accordance with the appropriate package insert, even if these expiration dates have passed.
- J. Maintain proper storage conditions during specimen shipping to ensure the integrity of the specimen. Specimen stability under shipping conditions other than those recommended has not been evaluated.
- K. Avoid cross-contamination during the specimen handling steps. Specimens can contain extremely high levels of virus or other organisms. Ensure that specimen containers do not come in contact with one another, and discard used materials without passing them over any open containers. Change gloves if they come in contact with specimens.
- L. Do not use the reagents and controls after the expiration date.
- M. Store assay components at the recommended storage condition. See *Reagent Storage and Handling Requirements* (page 6), and *Panther Fusion System Test Procedure* (page 11) for more information.

- N. Do not combine any assay reagents or fluids. Do not top off reagents or fluids; the Panther Fusion system verifies reagent levels.
- O. Avoid microbial and ribonuclease contamination of reagents.
- P. Quality control requirements must be performed in conformance with local, state, and/or federal regulations or accreditation requirements and your laboratory's standard quality control procedures. Reference to CLSI document C24-A3, *Statistical Quality Control for Quantitative Measurements: Principles and Definitions*: [Approved Guideline – Third Edition] or other published guidelines for general quality control is recommended. For further guidance on appropriate quality control practices, refer to 42 CFR 493.1205.
- Q. Do not use the assay cartridge if the storage pouch has lost its seal or if the assay cartridge foil is not intact. Contact Hologic if either occurs.
- R. Do not use the fluid packs if the foil seal is leaking. Contact Hologic if this occurs.
- S. Handle the assay cartridges with care. Do not drop or invert assay cartridges. Avoid prolonged exposure to ambient light.

	Panther Fusion Oil <i>Polydimethylsiloxane 100%</i> Warning H315 - Causes skin irritation H319 - Causes serious eye irritation
 	Panther Fusion Enhancer Reagent-S <i>Lithium Hydroxide Monohydrate 5-10%</i> Danger H302 - Harmful if swallowed H314 - Causes severe skin burns and eye damage P280 - Wear protective gloves/protective clothing/eye protection/face protection P260 - Do not breathe dust/fume/gas/mist/vapours/spray P303 + P361 + P353 - IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower P280 - Wear eye protection/ face protection P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing P310 - Immediately call a POISON CENTER or doctor/physician

Note: For information on any hazard and precautionary statements that may be associated with reagents, refer to the Safety Data Sheet Library at www.hologic.com/sds.

Reagent Storage and Handling Requirements

A. The following table provides storage and handling requirements for this assay.

Reagent	Unopened Storage	On Board/ Open Stability ¹	Opened Storage
Panther Fusion Flu A/B/RSV Assay Cartridge	2°C to 8°C	60 days	2°C to 8°C ²
Panther Fusion Capture Reagent-S (FCR-S)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Enhancer Reagent-S (FER-S)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Internal Control-S (IC-S)	2°C to 8°C	(In wFCR-S)	Not applicable
Panther Fusion Elution Buffer	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Oil	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Reconstitution Buffer I	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Flu A/B/RSV Positive Control	2°C to 8°C	Single use vial	Not applicable-single use
Panther Fusion Negative Control	2°C to 8°C	Single use vial	Not applicable-single use

When reagents are removed from the Panther Fusion system, return them immediately to their appropriate storage temperatures.

¹ On board stability starts at the time the reagent is placed on the Panther Fusion system for the Panther Fusion Flu A/B/RSV assay cartridge, FCR-S, FER-S and IC-S. On board stability starts for the Panther Fusion Reconstitution Buffer I, Panther Fusion Elution Buffer and Panther Fusion Oil Reagent when the reagent pack is first used.

² If removed from the Panther Fusion System, store the assay cartridge in an air-tight container with desiccant at the recommended storage temperature.

- B. Working Panther Fusion Capture Reagent-S and Panther Fusion Enhancer Reagent-S are stable for 60 days when capped and stored at 15°C to 30°C. Do not refrigerate.
- C. Discard any unused reagents that have surpassed their on board stability.
- D. Controls are stable until the date indicated on the vials.
- E. Avoid cross-contamination during reagent handling and storage.
- F. **Do not freeze reagents.**

Specimen Collection and Storage

Specimens - Clinical material collected from patient placed in an appropriate transport system. For the Panther Fusion Flu A/B/RSV assay, this includes NP swab specimens in viral transport medium (VTM).

Samples - Represents a more generic term to describe any material for testing on the Panther Fusion System including specimens, specimens transferred into a Panther Fusion Specimen Lysis Tube and controls.

Note: Handle all specimens as if they contain potentially infectious agents. Use Universal Precautions.

Note: Take care to avoid cross-contamination during specimen handling steps. For example, discard used material without passing over open tubes.

A. Specimen collection

Collect NP swab specimens according to standard technique using a polyester-, rayon-, or nylon-tipped swab. Immediately place the swab specimen into 3mL of VTM.

The following types of VTM were verified for use.

- Remel MicroTest M4, M4RT, M5 or M6 formulations
- Copan Universal Transport Medium
- BD Universal Viral Transport Medium

B. Specimen processing

1. Prior to testing on the Panther Fusion system, transfer specimen* to a Panther Fusion Specimen Lysis Tube.

- Transfer 500 µL of the NP swab specimens to a Panther Fusion Specimen Lysis Tube.

***Note:** When testing frozen specimen, allow specimen to reach room temperature prior to processing.

2. Storing specimens before testing

a. After collection, specimens can be stored at 2°C to 8°C up to 96 hours before transferred to the Panther Fusion Specimen Lysis Tube. Remaining specimen volumes can be stored at ≤-70°C.

b. Specimen in the Panther Fusion Specimen Lysis Tube may be stored under one of the following conditions:

- 15°C to 30°C up to 6 days or
- 2°C to 8°C up to 3 months.

Note: It is recommended that specimens transferred to the Panther Fusion Specimen Lysis Tube are stored capped and upright in a rack.

C. Samples on board the Panther Fusion system may be archived for additional testing at a later time.

D. Storing samples after testing

1. Samples that have been assayed should be stored upright in the rack under one of the following conditions:
 - 15°C to 30°C up to 6 days or
 - 2°C to 8°C up to 3 months.
2. The samples should be covered with a new, clean plastic film or foil barrier.
3. If assayed samples need to be frozen or shipped, remove the penetrable cap and place a new non-penetrable cap on the specimen tubes. If samples need to be shipped for testing at another facility, recommended temperatures must be maintained. Prior to uncapping previously tested and recapped samples, specimen transport tubes must be centrifuged for 5 minutes at 420 Relative Centrifugal Force (RCF) to bring all of the liquid down to the bottom of the tube. Avoid splashing and cross-contamination.

Specimen Transport

Maintain specimen storage conditions as described in the *Specimen Collection and Storage* section on page 7.

Note: *Specimens must be shipped in accordance with applicable national, international, and regional transportation regulations.*

Panther Fusion System

The Panther Fusion System is an integrated nucleic acid testing system that fully automates all steps necessary to perform various Panther Fusion assays from sample processing through amplification, detection, and data reduction.

Reagents and Materials Provided for the Panther Fusion Flu A/B/RSV Assay

Assay Packaging

Components ¹	Part No.	Storage
Panther Fusion Flu A/B/RSV Assay Cartridges 96 Tests Panther Fusion Flu A/B/RSV assay cartridge, 12 tests, 8 per box	PRD-04328	2°C to 8°C
Panther Fusion Internal Control-S 960 Tests Panther Fusion Internal Control-S tube, 4 per box	PRD-04332	2°C to 8°C
Panther Fusion Flu A/B/RSV Assay Controls Panther Fusion Flu A/B/RSV Positive Control tube, 5 per box Panther Fusion Negative Control tube, 5 per box	PRD-04336	2°C to 8°C
Panther Fusion Extraction Reagent-S 960 Tests Panther Fusion Capture Reagent-S bottle, 240 tests, 4 per box Panther Fusion Enhancer Reagent-S bottle, 240 tests, 4 per box	PRD-04331	15°C to 30°C
Panther Fusion Elution Buffer 2400 Tests Panther Fusion Elution Buffer pack, 1200 tests, 2 per box	PRD-04334	15°C to 30°C
Panther Fusion Reconstitution Buffer I 1920 Tests Panther Fusion Reconstitution Buffer I pack, 960 tests, 2 per box	PRD-04333	15°C to 30°C
Panther Fusion Oil Reagent 1920 Tests Panther Fusion Oil Reagent pack, 960 tests, 2 per box	PRD-04335	15°C to 30°C

¹ Components can also be ordered in the following bundles:

Panther Fusion Universal Fluids Kit, PRD-04430, contains 1 each Panther Fusion Oil and Panther Fusion Elution buffer.

Panther Fusion Assay Fluids I-S, PRD-04431, contains 2 Panther Fusion Extraction Reagents-S, 2 Panther Fusion Internal Control-S, and 1 Panther Fusion Reconstitution Buffer I.

Individually Packaged Items

Items	Part No.
Panther Fusion Specimen Lysis Tubes, 100 per bag	PRD-04339

Materials Required and Available Separately

Note: Materials available from Hologic have catalog numbers listed, unless otherwise specified.

Material	Cat. No.
Panther System	303095
Panther Fusion Module	ASY-09600
Aptima Assay Fluids Kit (Aptima Wash Solution, Aptima Buffer for Deactivation Fluid, and Aptima Oil Reagent)	303014 (1000 tests)
Multi-tube units (MTUs)	104772-02
Panther Waste Bag Kit	902731
Panther Waste Bin Cover	504405
Or Panther System Run Kit for Real Time Assays contains MTUs, waste bags, waste bin covers, and assay fluids	PRD-03455 (5000 tests)
Or Panther System Run Kit (when running TMA assays in parallel with real time-TMA assays) contains MTUs, waste bags, waste bin covers, auto detect*, and assay fluids	303096 (5000 tests)
Panther Fusion Tube Trays, 1008 tests, 18 trays per box	PRD-04000
Liquid Handling (LiHa) Disposable Tips, 1000 µL	10612513 (Tecan)
Aptima penetrable caps (optional)	105668
Replacement non-penetrable caps (optional)	103036A
Replacement extraction reagent bottle caps	CL0040
P1000 pipettor and tips with hydrophobic plugs	-
Bleach, 5% to 7% (0.7 M to 1.0 M) sodium hypochlorite solution Note: Mix one part bleach with one part deionized water to make diluted working bleach solution 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution.	-
Disposable powderless gloves	-

*Needed only for Panther Aptima TMA assays.

Panther Fusion System Test Procedure

Note: Refer to the Panther Fusion System Operator's Manual for additional procedural information.

A. Work Area Preparation

1. Wipe down work surfaces with 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution. Allow the sodium hypochlorite solution to contact surfaces for at least 1 minute and follow with a deionized (DI) water rinse. Do not allow the sodium hypochlorite solution to dry. Cover the bench surface with clean, plastic-backed absorbent laboratory bench covers.
2. Clean a separate work surface where samples will be prepared using the procedure described in step A.1.

B. Reagent Preparation

1. Remove the bottles of IC-S, FCR-S and FER-S from storage.
2. Open the bottles of IC-S, FCR-S and FER-S, and discard the caps. Open the TCR door on the upper bay of the Panther Fusion system.
3. Place the IC-S, FCR-S and FER-S bottles in the appropriate positions on the TCR carousel.
4. Close the TCR door.

Note: The Panther Fusion system adds the IC-S to the FCR-S. After the IC-S is added to the FCR-S, it is referred to as wFCR-S (working FCR-S). If the FCR-S and FER-S are removed from the system, use new caps and immediately store according to the proper storage conditions.

C. Specimen Handling

Note: Prepare specimens per the Specimen Processing instructions in the Specimen Collection and Storage section before loading specimens onto the Panther Fusion system.

1. **Do not vortex samples.**
2. Inspect sample tubes before loading into the rack. If a sample tube contains bubbles or has a lower volume than is typically observed, gently tap the bottom of the tube to bring contents to the bottom.

Note: To avoid a processing error, ensure adequate specimen volume is added to the Panther Fusion Specimen Lysis Tube. When 500 μ L of NP swab specimen is added to the Panther Fusion Specimen Lysis Tube, there is sufficient volume to perform 3 nucleic acid extractions.

D. System Preparation

For instructions on setting up the Panther Fusion system including loading samples, reagents, assay cartridges and universal fluids, refer to the Panther Fusion System Operator's Manual.

Procedural Notes

A. Controls

1. The Panther Fusion Flu A/B/RSV Positive Control and Panther Fusion Negative Control can be loaded in any rack position, in any Sample Bay lane on the Panther Fusion system.
2. Once the control tubes are pipetted and are processed for the Panther Fusion Flu A/B/RSV assay, they are active for up to 30 days (control frequency configured by an administrator) unless control results are invalid or a new assay cartridge lot is loaded.
3. Each control tube can be tested once.
4. Patient specimen pipetting begins when one of the following two conditions is met:
 - a. Valid results for the controls are registered on the system.
 - b. A pair of controls is currently in process on the system.

Quality Control

A run or specimen result may be invalidated by the Panther Fusion System if problems occur while performing the assay. Specimens with invalid results must be retested.

Negative and Positive Controls

To generate valid results, a set of assay controls must be tested. One replicate of the negative assay control and positive assay control must be tested each time a new lot of assay cartridges is loaded on the Panther Fusion system or when the current set of valid controls for an active cartridge lot have expired.

The Panther Fusion system is configured to require assay controls run at an administrator-specified interval of up to 30 days. Software on the Panther Fusion system alerts the operator when assay controls are required and does not start new tests until the assay controls are loaded and have started processing.

During processing, criteria for acceptance of the assay controls are automatically verified by the Panther Fusion system. To generate valid results, the assay controls must pass a series of validity checks performed by the Panther Fusion system.

If the assay controls pass all validity checks, they are considered valid for the administrator-specified time interval. When the time interval has passed, the assay controls are expired by the Panther Fusion system and requires a new set of assay controls be tested prior to starting any new samples.

If any one of the assay controls fails the validity checks, the Panther Fusion system automatically invalidates the affected samples and requires a new set of assay controls be tested prior to starting any new samples.

Internal Control

An internal control is added to each sample during the extraction process. During processing, the internal control acceptance criteria is automatically verified by the Panther Fusion system software. Detection of the internal control is not required for samples that are positive for Flu A, Flu B and/or RSV. The internal control must be detected in all samples that are negative for

Flu A, Flu B and RSV targets; samples that fail to meet that criteria will be reported as Invalid. Each sample with an Invalid result must be retested.

The Panther Fusion system is designed to accurately verify processes when procedures are performed following the instructions provided in this package insert and the *Panther Fusion System Operator's Manual*.

Interpretation of Results

The Panther Fusion system automatically determines the test results for samples and controls. Results for Flu A, Flu B and RSV detection are reported separately. A test result may be negative, positive, or invalid.

Table1 shows the possible results reported in a valid run with result interpretations.

Table 1: Result Interpretation

Flu A Result	Flu B Result	RSV Result	IC Result	Interpretation
Neg	Neg	Neg	Valid	Flu A, Flu B, and RSV not detected.
POS	Neg	Neg	Valid	Flu A detected. Flu B and RSV not detected.
Neg	POS	Neg	Valid	Flu B detected. Flu A and RSV not detected.
Neg	Neg	POS	Valid	RSV detected. Flu A and Flu B not detected.
POS	POS	Neg	Valid	Flu A and Flu B detected. RSV not detected.
Neg	POS	POS	Valid	Flu B and RSV detected. Flu A not detected.
POS	Neg	POS	Valid	Flu A and RSV detected. Flu B not detected.
POS	POS	POS	Valid	Flu A, Flu B, and RSV detected. Triple infections are rare. Retest to confirm result.
Invalid	Invalid	Invalid	Invalid	Invalid. There was an error in the generation of the result; retest sample.

Note: POS result will be accompanied by cycle threshold (Ct) values.

Limitations

- A. Use of this assay is limited to personnel who are trained in the procedure. Failure to follow these instructions may result in erroneous results.
- B. Reliable results are dependent on adequate specimen collection, transport, storage, and processing.
- C. Avoid contamination by adhering to good laboratory practices and to the procedures specified in this package insert.
- D. Negative results do not preclude influenza A virus, influenza B virus, or RSV infections and should not be used as the sole basis for treatment or other management decisions.
- E. This test does not differentiate influenza A subtypes (i.e. H1N1, H3N2) or RSV subgroups (i.e., A or B); additional testing is required to differentiate any specific influenza A subtypes or strains or specific RSV subgroups, in consultation with local public health departments.
- F. A positive result indicates the detection of nucleic acid from the relevant virus. Nucleic acid may persist even after the virus is no longer viable.

Panther Fusion System Assay Performance

Clinical Performance

Retrospectively collected NP swab specimens from patients in the US with reference test results were used for evaluation. The results are shown in tables 2, 3, and 4.

For NP swab specimens, 500 microliter (µL) was diluted into a Specimen Lysis Tube containing 780 µL of Specimen Transport Media (STM) and a single replicate was tested with the Panther Fusion Flu A/B/RSV assay. The result was compared to an FDA approved nucleic acid test (NAT) result. The sensitivity and specificity for the detection of Flu A, Flu B, and RSV nucleic acid was determined.

A total of 716 NP swab specimens were tested with the Panther Fusion Flu A/B/RSV assay and with Luminex xTAG® Respiratory Viral Panel or Luminex xTAG® Respiratory Viral Panel FAST v2 or GenMark Dx eSensor Respiratory Viral Panel. Sensitivity and specificity for detection of Flu A, Flu B and RSV are shown.

Table 2: Flu A Results

Specimen Type	N	Flu A+		Flu A-		Sensitivity or Positive Agreement 95% CI	Specificity or Negative Agreement 95% CI	Overall Agreement 95% CI
		Fusion Flu A +	Fusion Flu A -	Fusion Flu A +	Fusion Flu A -			
Nasopharyngeal Swab	716	331	4*	4**	377	98.8% 97.0 - 99.5%	99.0% 97.3 - 99.6%	98.9% 97.8 - 99.4%

* Two out of 4 discordant specimens tested with an in-house developed and validated RT-PCR assay. Flu A was not detected in both specimens. Untested discordant specimens had insufficient volumes.

** All 4 discordant specimens were tested with an in-house developed and validated RT-PCR assay. Flu A was detected in 3 out of 4 specimens.

Table 3: Flu B Results

Specimen Type	N	Flu B+		Flu B-		Sensitivity or Positive Agreement 95% CI	Specificity or Negative Agreement 95% CI	Overall Agreement 95% CI
		Fusion Flu B +	Fusion Flu B -	Fusion Flu B +	Fusion Flu B -			
Nasopharyngeal Swab	716	74	0	1*	641	100.0% 95.1 - 100.0%	99.8% 99.1 - 100.0%	99.9% 99.2 - 100.0%

* Flu B was detected when tested with an in-house developed and validated RT-PCR assay.

Table 4: RSV Results

Specimen Type	N	RSV+		RSV-		Sensitivity or Positive Agreement 95% CI	Specificity or Negative Agreement 95% CI	Overall Agreement 95% CI
		Fusion RSV +	Fusion RSV -	Fusion RSV +	Fusion FSV -			
Nasopharyngeal Swab	716	305	2*	4**	405	99.3% 97.7 - 99.8%	99.0% 97.5 - 99.6%	99.2% 98.2 - 99.6%

*Both discordant specimens were tested with an in-house developed and validated RT-PCR assay. RSV was not detected.

** Two out of 4 discordant specimens were tested with an in-house developed and validated RT-PCR assay. RSV was detected in both specimens. Untested discordant specimens had insufficient volumes.

Analytical Sensitivity

The analytical sensitivity (limit of detection or LoD) of the Panther Fusion Flu A/B/RSV assay was determined by testing pooled Flu A/B/RSV negative clinical specimens spiked with the following virus cultures at various concentrations: 4 Flu A strains, 2 Flu B strains, 1 strain each for RSV A and RSV B. Twelve replicates were tested with each of the three reagent lots for a combined total of 36 replicates. Target specific LoD concentrations were verified by testing an additional 20 replicates with one reagent lot. Analytical sensitivity (LoD) is defined as the lowest concentration at which $\geq 95\%$ of all replicates tested positive, as summarized in table 5.

Table 5: NP Swab Sensitivity

Viral Strain	LoD Concentration
Influenza A/California/07/2009 (H1N1)	$1 \times 10^{-1.0}$ TCID ₅₀ /mL
Influenza A/Massachusetts/15/13 (H1N1)	$1 \times 10^{-1.5}$ TCID ₅₀ /mL
Influenza A/Switzerland/9715293/2013 (H3N2)	$1 \times 10^{-1.5}$ TCID ₅₀ /mL
Influenza A/Victoria/361/2011 (H3N2)	$1 \times 10^{-1.5}$ TCID ₅₀ /mL
Influenza B/Brisbane/33/08	$1 \times 10^{-0.5}$ TCID ₅₀ /mL
Influenza B/Massachusetts/02/2012	$1 \times 10^{-2.0}$ TCID ₅₀ /mL
RSV A	$1 \times 10^{0.5}$ TCID ₅₀ /mL
RSV B	$1 \times 10^{0.0}$ TCID ₅₀ /mL

Reactivity

The reactivity of the Panther Fusion assay was evaluated against multiple strains of Influenza A, Influenza B, and Respiratory Syncytial Viruses. Viral strains were tested in triplicates with each of the three reagent lots for a combined total of 9 replicates. Viruses present at concentrations below those tested for Reactivity may not be detected by the Panther Fusion Flu A/B/RSV assay.

Table 6: Analytical Reactivity (inclusivity) Test Summary

Description	Type	Concentration	Flu A	Flu B	RSV
A/Aichi/2/1968	Influenza A/H3N2	1x10 ² CEID ₅₀ /mL	+	-	-
A/Brazil/02/1999	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Brazil/1137/1999	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Brisbane/59/2007	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/California/07/2009	Influenza A/H1N1	1x10 ⁻¹ TCID ₅₀ /mL	+	-	-
A/Costa Rica/07/1999	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Denver/1/57	Influenza A/H1N1	1x10 ² CEID ₅₀ /mL	+	-	-
A/Dominican Republic/7293/13	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Fujian/156/2000	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Georgia/F32551/12 2009	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Hawaii/15/2001	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Henan/8/2005	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Hiroshima/52/2005	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Hong Kong/218/2006	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Hong Kong/4801/2014	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Hong Kong/486/97 RNA	Influenza A/H5N1	16.4 ng/mL	+	-	-
A/Hong Kong/8/1968	Influenza A/H3N2	1x10 ² CEID ₅₀ /mL	+	-	-
A/Indiana/08/2011	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Japan/305/1957	Influenza A/H2N2	0.003 ug/mL	+	-	-
A/Jiangxi/160/2005	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Kentucky/2/2006	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Malaya/302/54	Influenza A/H1N1	1x10 ² CEID ₅₀ /mL	+	-	-
A/Mexico/4108/2009	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Minnesota/11/2010	Influenza A/H3N2	36 ng/mL	+	-	-
A/New Jersey/8/1976	Influenza A/H1N1	1x10 ³ TCID ₅₀ /mL	+	-	-
A/Ohio/09SW1477/2009	Influenza A/H1N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Perth/16/2009	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Port Chalmers/1/1973	Influenza A/H3N2	1x10 ² TCID ₅₀ /mL	+	-	-
A/Puerto Rico/8/34	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Solomon Islands/03/2009	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Switzerland/9715293/2013	Influenza A/H3N2	1x10 ^{-1.5} TCID ₅₀ /mL	+	-	-
A/Taiwan/42/2006	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Victoria/3/1975	Influenza A/H3N2	1x10 ² CEID ₅₀ /mL	+	-	-
A/Vietnam/1203 RNA	Influenza A/H5N1	0.27 ug/mL	+	-	-
A/WS/33	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
B/Brisbane/60/2008	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Florida/2/2006 (Yamagata lineage)	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Florida/7/2004	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Hawaii/11/2005	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Hawaii/33/2004	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-

Table 6: Analytical Reactivity (inclusivity) Test Summary (continued)

Description	Type	Concentration	Flu A	Flu B	RSV
B/Lee/40	Influenza B	1x10 ² CEID ₅₀ /mL	-	+	-
B/Michigan/2/2006	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Ohio/1/2005	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Panama/45/90	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/Phuket/3073/2013 (Victoria Lineage)	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
B/St. Petersburg/04/2006	Influenza B	1x10 ² TCID ₅₀ /mL	-	+	-
RSV A/A2	RSV	1x10 ² TCID ₅₀ /mL	-	-	+
RSV A/Long	RSV	1x10 ² TCID ₅₀ /mL	-	-	+
RSV A/Vero	RSV	1x10 ² CEID ₅₀ /mL	-	-	+
RSV B/9320	RSV	1x10 ² TCID ₅₀ /mL	-	-	+
RSV B/Wash/18537/62	RSV	2x10 ² TCID ₅₀ /mL	-	-	+

Table 7: Additional Analytical Reactivity (inclusivity) Test Summary

Description	Type	Concentration	Flu A	Flu B	RSV
A/Chicken/Germany/N/49	Influenza A/H10N7	68 ng/mL	+	-	-
A/Duck/Alberta/35/76	Influenza A/H1N1	1 ng/mL	+	-	-
A/Duck/Chabarovsk/1610/1972	Influenza A/H3N8	1 ng/mL	+	-	-
A/Duck/Czechoslovakia/1956	Influenza A/H4N6	2.6 ng/mL	+	-	-
A/Duck/Memphis/546/1974	Influenza A/H11N9	8 ng/mL	+	-	-
A/Duck/Pennsylvania/10218/1984	Influenza A/H5N2	3 ng/mL	+	-	-
A/Duck/Singapore/645/97	Influenza A/H5N3	2 ng/mL	+	-	-
A/Duck/Ukraine/1963	Influenza A/H3N8	3 ng/mL	+	-	-
A/gyrfalcon/Washington/41088-6/2014	Influenza A/H5N8	1x10 ³ TCID ₅₀ /mL	+	-	-
A/Northern pintail/Washington/40964/2014	Influenza A/H5N2	1x10 ³ TCID ₅₀ /mL	+	-	-
A/Swine/ NY/01/2009	Influenza A/H1N1	1x10 ² TCID ₅₀ /mL	+	-	-
A/Swine/Iowa/2006	Influenza A/H1N1	1x10 ² CEID ₅₀ /mL	+	-	-
A/Turkey/Massachusetts/3740/1965	Influenza A/H6N2	1 ng/mL	+	-	-
A/Turkey/Ontario/6118/1968	Influenza A/H8N4	2 ng/mL	+	-	-
A/Turkey/Wisconsin/1/1966	Influenza A/H9N2	23 ng/mL	+	-	-

Analytical Specificity

The analytical specificity of the Panther Fusion Flu A/B/RSV assay was evaluated by testing a panel of 52 organisms, consisting of 25 viral, 26 bacterial, and 1 yeast strain representing common respiratory pathogens or flora commonly present in respiratory tract. Bacteria and yeast were tested at concentrations of 10⁵ to 10⁸ CFU/mL or IFU/mL, except where noted. Viruses were tested at concentrations of 10³ to 10⁷ TCID₅₀/mL.

Analytical specificity of the Panther Fusion Flu A/B/RSV assay was 100% for Flu A, Flu B, and RSV.

Table 8: Specificity Results

Organism	Concentration	Flu A	Flu B	RSV
Adenovirus 1	1x10 ⁵ TCID ₅₀ /mL	-	-	-
Adenovirus 7a	1x10 ⁵ TCID ₅₀ /mL	-	-	-
<i>Bordetella bronchiseptica</i>	1x10 ⁷ CFU/ml	-	-	-
<i>Bordetella pertussis</i>	1x10 ⁸ CFU/mL	-	-	-
<i>Candida albicans</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Chlamydia trachomatis</i>	1x10 ⁵ CFU/mL	-	-	-
<i>Chlamydophila pneumoniae</i> (formerly <i>Chlamydia pneumoniae</i>)	1x10 ⁵ IFU/mL	-	-	-
CMV Strain AD 169	1x10 ⁴ TCID ₅₀ /mL	-	-	-
Coronavirus 229E	1x10 ⁴ TCID ₅₀ /mL	-	-	-
<i>Corynebacterium diphtheria</i>	1x10 ⁷ CFU/mL	-	-	-
Coxsackie B4	1x10 ⁶ TCID ₅₀ /mL	-	-	-
Coxsackie B5/10/2006	1x10 ⁵ TCID ₅₀ /mL	-	-	-
<i>E. coli</i>	1x10 ⁷ CFU/mL	-	-	-
EBV	1x10 ⁷ TCID ₅₀ /mL	-	-	-
Echovirus 2	1x10 ⁴ TCID ₅₀ /mL	-	-	-
Echovirus 3	1x10 ⁵ TCID ₅₀ /mL	-	-	-
Echovirus 6	1x10 ⁴ TCID ₅₀ /mL	-	-	-
Echovirus 11	1x10 ⁵ TCID ₅₀ /mL	-	-	-
Enterovirus 68	1x10 ⁵ TCID ₅₀ /mL	-	-	-
Enterovirus 70	1x10 ⁴ TCID ₅₀ /mL	-	-	-
<i>Haemophilus Influenzae</i>	1x10 ⁷ CFU/mL	-	-	-
hMPV Subtype A2	1x10 ⁶ TCID ₅₀ /mL	-	-	-
HPIV-1	1x10 ⁴ TCID ₅₀ /mL	-	-	-
HPIV-2	1x10 ⁵ TCID ₅₀ /mL	-	-	-
HPIV-3	1x10 ⁵ TCID ₅₀ /mL	-	-	-
HPIV-4	1x10 ⁴ TCID ₅₀ /mL	-	-	-
HSV-1 Macintyre Strain	1x10 ⁵ TCID ₅₀ /mL	-	-	-
HSV-2 Type 2G Strain	1x10 ⁵ TCID ₅₀ /mL	-	-	-
<i>Klebsiella pneumonia</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Lactobacillus plantarum</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Legionella pneumophila</i>	1x10 ⁷ CFU/mL	-	-	-
Measles/7/2000	1x10 ⁵ TCID ₅₀ /mL	-	-	-
<i>Moraxella catarrhalis</i>	1x10 ⁶ CFU/mL	-	-	-

Table 8: Specificity Results (continued)

Organism	Concentration	Flu A	Flu B	RSV
Mumps virus	1x10 ⁴ TCID ₅₀ /mL	-	-	-
<i>Mycobacterium intracellulare</i>	1x10 ¹⁰ rRNA copies/mL	-	-	-
<i>Mycobacterium tuberculosis</i>	1x10 ¹⁰ rRNA copies/mL	-	-	-
<i>Mycoplasma pneumoniae</i>	1x10 ⁶ CFU/mL	-	-	-
<i>Neisseria gonorrhoea</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Neisseria meningitides</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Neisseria mucosa</i>	1x10 ⁷ CFU/mL	-	-	-
Polio virus	1x10 ⁶ TCID ₅₀ /mL	-	-	-
<i>Proteus mirabilis</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Proteus vulgaris</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Pseudomonas aeruginosa</i>	1x10 ⁷ CFU/mL	-	-	-
Rhinovirus 1A	1x10 ⁵ TCID ₅₀ /mL	-	-	-
<i>Staphylococcus aureus</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Staphylococcus epidermidis</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Streptococcus pneumoniae</i>	1x10 ⁶ CFU/mL	-	-	-
<i>Streptococcus pyogenes</i>	1x10 ⁷ CFU/mL	-	-	-
<i>Streptococcus salivarius</i>	1x10 ⁶ CFU/mL	-	-	-
<i>Tatlockia micdadei</i> (formerly <i>Legionella micdadei</i>)	1x10 ⁷ CFU/mL	-	-	-
Varicella Zoster Virus	1x10 ³ TCID ₅₀ /mL	-	-	-

Competitive Interference

Competitive Interference of the Panther Fusion Flu A/B/RSV assay was evaluated using a simulated clinical matrix with pairs of target viruses at two different concentrations. One of the concentrations was near the Limit of Detection (3 - 5X LoD) while the other concentration was high (1000X LoD). The presence of two viruses at varying concentrations in a single sample had no effect on the analytical sensitivity (100% detection for both targets) at the concentration noted in Table 9.

Table 9: Competitive Interference

Condition	Target 1		Target 2		Flu A	Flu B	RSV
	Description	Concentration	Description	Concentration			
1	FLU A	3X LoD	RSV	1000X LoD	+	-	+
2	FLU A	3X LoD	FLU B	1000X LoD	+	+	-
3*	FLU B	5X LoD	FLU A	1000X LoD	+	+	-
4	FLU B	3X LoD	RSV	1000X LoD	-	+	+
5	RSV	3X LoD	FLU A	1000X LoD	+	-	+
6	RSV	3X LoD	FLU B	1000X LoD	-	+	+

*When this combination was tested with Flu B at 3X LoD, Flu B detection rate was 92.3%.

Interference

Mucin, whole blood and other potentially interfering substances (medications and over-the-counter or OTC products) that may be present in the samples were evaluated in the Panther Fusion Flu A/B/RSV assay. Clinically relevant amount of the potentially interfering substances were added to simulated clinical matrix and tested unspiked or spiked with cultured Flu A, Flu B and RSV at their respective 3X LoD concentrations. The substances consisted of nasal sprays (liquid and powder), ingestible pills, lozenges, injectable and endogenous substances, as shown in Table 10.

All of the substances tested were found to have no impact on the performance of the Panther Fusion Flu A/B/RSV assay.

Table 10: Potentially Interfering Substances

Type	Substance Name	Active Ingredient(s)	Concentration
Endogenous	Mucin	Purified mucin protein	60 µg/mL
	Human blood	Blood	2% v/v
Nasal sprays or drops	Neo-Syneprine®	Phenylephrine	15% v/v
	Anefrin	Oxymetazoline	15% v/v
	Saline	Sodium chloride	15% v/v
	Ventolin® HFA	Albuterol	15% v/v
Nasal corticosteroids	QVAR®, Beconase AQ	Beclomethasone	5% v/v
	Dexacort	Dexamethasone	5% v/v
	AEROSPAN®	Flunisolide	5% v/v
	Nasacort	Triamcinolone	5% v/v
	Rhinocort	Budesonide	5% v/v
	Nasonex	Mometasone	5% v/v
Nasal gel	Zicam® (Allergy Relief)	Luffa operculata, Galphimia, Glauca, Histaminum hydrochloricum, Sulfur	5% v/v
Throat lozenges	Chloraseptic Throat Lozenges	Benzocaine Menthol	0.63 mg/mL
Anti-viral drugs	Relenza®	Zanamivir	3.3 mg/mL
	TamiFlu	Oseltamivir	25 mg/mL
	Rebitol	Ribavirin	20 mg/mL
Antibiotic, nasal ointment	Bactroban cream	Mupirocin	10 mg/mL
Antibiotic, systemic	Tobramycin	Tobramycin	4.0 µg/mL

Carry-Over/Contamination

The carry-over/cross-contamination study was performed with negative samples alternately placed between high positive samples and tested. High positive samples were prepared by spiking (over 10,000X LoD). A total of nine separate runs with negative samples and positive samples placed in a checkerboard pattern were tested over three different instruments for a combined total of 449 positive and 449 negative samples. The carry-over rate was 0.4%.

Assay Precision

Panther Fusion Flu A/B/RSV assay precision was evaluated with a 7-member panel. The panel was tested by three operators on two separate runs per day, using three reagent lots on three Panther Fusion systems over 45 days.

The panel members are described in Table 11, along with a summary of the agreement with expected results for each targets. Table 12 presents the mean and variability analysis between instruments, between reagent lots, between operators, between days, between runs and within runs, and overall (total) for Ct.

Table 11: Percent Agreement to the Expected Result

Target	Panel Member	% Positive	% Agreement (95% CI)
Flu A	Flu A 3x LoD	100.0% (162/162)	100.0% (97.7 - 100%)
	Flu A 1x LoD	100.0% (162/162)	100.0% (97.7 - 100%)
	Flu A 0.01x LoD	8.6% (14/162)	91.4% (86.0 - 94.8%)
	Negative	0.0% (0/162)	100.0% (97.7 - 100%)
Flu B	Flu B 3x LoD	100.0% (162/162)	100.0% (97.7 - 100%)
	Flu B 1x LoD	94.4% (153/162)	94.4% (89.8 - 97.0%)
	Flu B 0.01x LoD	4.3% (7/162)	95.7% (91.4 - 97.9%)
	Negative	0.6% (1/162)	99.4% (96.6 - 99.9%)
RSV	RSV 3x LoD	100.0% (162/162)	100.0% (97.7 - 100%)
	RSV 1x LoD	99.4% (161/162)	99.4% (96.6 - 99.9%)
	RSV 0.01x LoD	4.9% (8/162)	95.1% (90.6 - 97.5%)
	Negative	0.0% (0/162)	100.0% (97.7 - 100%)

Table 12: Signal Variability

Target	Panel Member	Mean Ct	Between Instrument		Between Reagent Lots		Between Operators		Between Days		Between Runs		Within Runs		Total	
			SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)
Flu A	Flu A 3x LoD	35.0	0.1	0.3	0.2	0.5	0.0	0.0	0.0	0.0	0.2	0.6	0.7	2.1	0.8	2.4
	Flu A 1x LoD	35.3	0.0	0.1	0.1	0.5	0.0	0.0	0.0	0.0	0.2	0.6	0.8	2.4	0.9	2.5
	Flu A 0.01x LoD	38.1	0.3	0.9	0.2	0.6	0.3	0.9	0.0	0.0	0.0	0.0	0.9	2.3	1.0	2.8
Flu B	Flu B 3x LoD	36.5	0.0	0.1	0.1	0.5	0.0	0.0	0.0	0.0	0.1	0.3	0.7	1.9	0.7	2.0
	Flu B 1x LoD	38.0	0.2	0.5	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.4	0.8	2.1	0.8	2.2
	Flu B 0.01x LoD	39.4	0.3	1.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.3	0.9	0.5	1.3
RSV	RSV 3x LoD	36.2	0.2	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.3	3.5	1.3	3.6
	RSV 1x LoD	38.2	0.3	0.8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	1.6	4.2	1.6	4.3
	RSV 0.01x LoD	40.7	0.0	0.0	0.0	0.0	0.2	0.6	0.4	1.0	0.0	0.0	0.2	0.5	0.5	1.3
IC	Negative	33.1	0.1	0.3	0.2	0.6	0.0	0.0	0.1	0.3	0.2	0.6	0.3	1.1	0.5	1.5

Bibliography

1. Centers for Disease Control and Prevention. National Respiratory and Enteric Virus Surveillance System. Centers for Disease Control and Prevention Web site. <http://www.cdc.gov/surveillance/nrevss/>. Accessed October, 2015.
2. Kahn, J.S. 2006. Epidemiology of human metapneumovirus. *Clin. Microbiol. Rev.* 19:546-557.
3. Couch, R.B. and Kasel, J.A. 1995. Influenza in *Diagnostic Procedures for Viral, Rickettsial, and Chlamydial Infections*. 7th Edition. 431-446.
4. Harper, S.A., Fukuda, K., Uyeki, T.M., Cox, N.J., and Bridges, C.B. 2005. Prevention and Control of Influenza. *MMWR*. 54(RR08):1-40.
5. World Health Organization. Influenza (Seasonal) Fact Sheet N° 211 March 2014. Centers for Disease Control and Prevention Web site. <http://www.who.int/mediacentre/factsheets/fs211/en/>. Accessed October 2015.
6. Centers for Disease Control and Prevention. Respiratory Syncytial Virus Circulation in the United States, July 2012-June 2014 *MMWR* 2014;62:141-4 Centers for Disease Control and Prevention Web site. www.cdc.gov/rsv/research/us-surveillance.html. Accessed October 2015.



Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA

Hologic N.V.
Da Vincilaan 5
1930 Zaventem
Belgium

Customer Support: +1 800 442 9892
customersupport@hologic.com

Technical Support: +1 888 484 4747
molecularsupport@hologic.com

For more contact information visit www.hologic.com.

Hologic and Panther Fusion are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at www.hologic.com/patents.

©2017-2018 Hologic, Inc. All rights reserved.

AW-16162-001 Rev. 003
2018-11