



Date: 11/26/2019

Author: Software Engineering

System: Apex

Subsystem: Microsoft

Windows OS 10

Subject: Microsoft Windows 10 OS Patch release September 2019

Purpose:

To release the list of validated Microsoft critical patches on the Apex systems.

Note: If you are using the following versions of APEX, if you are required to update to the latest Hologic validated Windows updates, you must ensure your system(s) have been patched with the "Sentinel Run-Time 7.80.0.0 " which can be located in the Windows Control Panel "Program and Features". If needed you can install the patch by ordering "APEX Service Dongle Patch Release" (ASY-12610). Please contact Hologic Support if you have any questions.

System Type	Software Versions
Horizon	APEX v5.6.0.7
Workstation	APEX v5.6.0.7

List of patches/KB's validated for September 2019 on APEX System:

KB4516115 KB4516044 KB4512574 KB4494175
KB4132216 KB2565063 KB2467173