

Compression Thickness Indicator Test Accuracy

Place this document with your Hologic® Selenia® Dimensions® *Quality Control Manual* for future reference.

This notice is to inform Hologic® Selenia® Dimensions® Digital Mammography users that the following note is added to the Selenia Dimensions *Quality Control Manual*, MAN-01965. This note is added to Chapter 3, *Quality Control Activities for the Medical Physicist*, Section 3.1, *Mammographic Unit Assembly Evaluation*, immediately following paragraph 3.1.6, *Recommended Performance Criteria and Corrective Action*.



Note

Accuracy of the compression breast thickness scale shall be conducted according to the *Compression Thickness Indicator* test procedure listed in this *Quality Control Manual* in Chapter 4, *Quality Control Activities for the Radiologic Technologist*, Section 4.7, *Compression Thickness Indicator*, paragraph 4.7.4, *Test Procedure*.
