

RoHS Declaration

Selenia Dimensions/3Dimensions Mammography System
Affirm Breast Biopsy Guidance System

Manufacturer's Name: Hologic, Inc.
Manufacturer's Address: 36 Apple Ridge Road
Danbury, CT 06810 USA

Type of Equipment: X-Ray Mammography
Stereotactic Biopsy Systems

Model Designation: Selenia® Dimensions® Mammography System
3Dimensions™ Mammography System
Affirm® Breast Biopsy Guidance System

The following exemptions defined in Directive 2011/65/EU (RoHS II) and 2015/863/EU are applied to the Selenia Dimensions/3Dimensions Mammography System and the Affirm Breast Biopsy Guidance System.

Annex	Exemption
IV-23	Lead as an alloying element for bearings and wear surfaces in medical equipment exposed to ionizing radiation.
III -7(c)-I	Electrical and electronic components containing lead in a glass or ceramic other than dielectric ceramic in capacitors, for example, piezoelectronic devices, or in a glass or ceramic matrix compound
III- 6(a)	Lead as an alloying element in steel for machining purposes and in galvanized steel containing up to 0.35% lead by weight
III- 6(c)	Copper alloy containing up to 4% lead by weight
IV-5	Lead in shielding for ionizing radiation