

**SARS-CoV-2/Flu A/B/RSV Assay (Panther Fusion™ System)**

For *in vitro* diagnostic use.

For U.S. Export only.

**CONTENTS**

**General Information** ..... 2

    Intended Use ..... 2

    Summary and Explanation of the Test ..... 2

    Principles of the Procedure ..... 3

    Warnings and Precautions ..... 4

    Reagent Storage and Handling Requirements ..... 7

    Specimen Collection and Storage ..... 8

    Specimen Transport ..... 9

**Panther Fusion System** ..... 10

    Reagents and Materials Provided for Panther Fusion SARS-CoV-2/Flu A/B/RSV Assay 10

    Materials Required and Available Separately ..... 11

    Panther Fusion System Test Procedure ..... 12

    Procedural Notes ..... 13

**Quality Control** ..... 14

**Interpretation of Results** ..... 14

**Limitations** ..... 16

**SARS-CoV-2/Flu A/B/RSV Assay Performance** ..... 17

**Bibliography** ..... 27

## General Information

### Intended Use

The Panther Fusion™ SARS-CoV-2/Flu A/B/RSV assay is a fully automated multiplexed real-time RT-PCR test intended for the qualitative detection and differentiation of RNA from SARS-CoV-2 virus, influenza A virus (Flu A), influenza B virus (Flu B) and respiratory syncytial virus (RSV) isolated and purified from nasopharyngeal (NP) swab specimens obtained from individuals exhibiting signs and symptoms of a respiratory tract infection. Clinical signs and symptoms of respiratory viral infection due to SARS-CoV-2, influenza and RSV can be similar. This assay is intended to aid in the differential diagnosis of SARS-CoV-2, influenza A virus, influenza B virus and RSV infections in humans and is not intended to detect influenza C virus infections.

Negative results do not preclude SARS-CoV-2, influenza A virus, influenza B virus or RSV infections and should not be used as the sole basis for treatment or other management decisions. This assay is designed for use on the Panther Fusion system.

### Summary and Explanation of the Test

Respiratory viruses are responsible for a wide range of acute respiratory tract infections including the common cold, influenza (flu), RSV infection, COVID-19 and croup and represent the most common cause of acute illness in the United States. Some symptoms of COVID-19, flu and RSV are similar making diagnosis based on symptoms alone virtually impossible.<sup>1,2</sup>

Disease severity of flu and RSV can be especially high in the young, the immunocompromised, and elderly patients. Accurate and timely diagnosis of the cause of respiratory tract infections has many benefits. They include improved treatment of the patient by ensuring appropriate antiviral treatment (e.g. oseltamivir for influenza),<sup>3</sup> decreasing the overall cost of care, reducing the potential for further development of antimicrobial resistance due to excessive and inappropriate use of antibiotics,<sup>4</sup> assisting infection control personnel in providing appropriate measures to minimize nosocomial spread, and providing valued information to public health authorities regarding which viruses are circulating in the community.<sup>5</sup>

Influenza is an acute respiratory illness caused by infection with the influenza virus, primarily types A and B.<sup>6</sup> Influenza A viruses are further categorized into subtypes based on the two major surface protein antigens: hemagglutinin (H) and neuraminidase (N).<sup>7</sup> Influenza B viruses are not categorized into subtypes.<sup>7</sup> Influenza viruses continuously undergo genetic changes including drift (random mutation) and variation (genomic reassortment), generating new strains of virus each year, leaving the human population vulnerable to these seasonal changes. Epidemics occur yearly (typically in winter) and while both types A and B circulate in the population, type A is usually dominant. Transmission of influenza is primarily via airborne droplet (coughing or sneezing). Symptoms arise on average 1 to 2 days post-exposure and include fever, chills, headache, malaise, cough, and coryza.

Complications due to influenza include pneumonia causing increased morbidity and mortality in pediatric, elderly and immunocompromised populations. Influenza occurs globally with an annual attack rate estimated at 5%–10% in adults and 20%–30% in children. Illnesses can result in hospitalization and death mainly among high-risk groups (the very young, elderly or chronically ill). Worldwide, these annual epidemics are estimated to result in about 3 to 5 million cases of severe illness, and about 250,000 to 500,000 deaths.<sup>8</sup>

Respiratory syncytial virus (RSV) is a leading cause of respiratory infections in infants and children. There are 2 types of RSV (A and B) based on antigenic and surface protein variations. Most yearly epidemics (typically during winter) contain a mix of type A and B viruses, but one subgroup can dominate during a season. RSV infection can cause severe respiratory illness among all ages but is more prevalent in pediatric, elderly and immunocompromised populations. Annually in the United States, RSV infection has been associated with an estimated 58,000 hospitalizations and 2.1 million outpatient visits among children younger than 5 years, and 177,000 hospitalizations and 14,000 deaths among adults older than 65 years.<sup>9</sup>

Coronaviruses are a large family of viruses which may cause illness in animals or humans. In humans, several coronaviruses are known to cause respiratory infections ranging from the common cold to more severe diseases such as Middle East Respiratory Syndrome (MERS) and Severe Acute Respiratory Syndrome (SARS). The most recently discovered coronavirus, SARS-CoV-2, causes the associated coronavirus disease COVID-19. This new virus and disease were unknown before outbreak in Wuhan, China, in December 2019.<sup>9</sup>

People with COVID-19 have had a wide range of symptoms reported, ranging from mild symptoms to severe illness. Symptoms may appear 2-14 days after exposure to the virus. People with COVID-19 may exhibit fever or chills, cough, shortness of breath or difficulty breathing, fatigue, muscle or body aches, headache, new loss of taste or smell, sore throat, congestion or runny nose, nausea or vomiting, and/or diarrhea.<sup>10</sup> On March 11, 2020, the COVID-19 outbreak was characterized as a pandemic by the World Health Organization (WHO).<sup>11</sup>

## Principles of the Procedure

The Panther Fusion SARS-CoV-2/Flu A/B/RSV assay involves the following steps: sample lysis, nucleic acid capture and elution transfer, and multiplex RT-PCR when analytes are simultaneously amplified, detected and differentiated. Nucleic acid capture and elution takes place in a single tube on the Panther Fusion system. The eluate is transferred to the Panther Fusion system reaction tube containing the assay reagents. Multiplex RT-PCR is then performed for the eluted nucleic acid on the Panther Fusion system.

**Nucleic acid capture and elution:** Prior to processing and testing on the Panther Fusion system, specimens are transferred to a Specimen Lysis Tube containing specimen transport media (STM). STM lyses the cells, releases target nucleic acid, and protects them from degradation during storage.

The Internal Control-S (IC-S) is added to each test specimen and controls via the working Panther Fusion Capture Reagent-S (wFCR-S). The IC-S in the reagent monitors specimen processing, amplification and detection.

Capture oligonucleotides hybridize to nucleic acid in the test specimen. Hybridized nucleic acid is then separated from the specimen in a magnetic field.

Wash steps remove extraneous components from the reaction tube. The elution step elutes purified nucleic acid. During the nucleic acid capture and elution step, total nucleic acid is isolated from specimens.

**Elution transfer and RT-PCR:** During the elution transfer step, eluted nucleic acid is transferred to a Panther Fusion reaction tube already containing oil and reconstituted mastermix.

Target amplification occurs via RT-PCR. A reverse transcriptase generates a DNA copy of the target sequence. Target specific forward and reverse primers and probes then amplify targets while simultaneously detecting and discriminating multiple target types via multiplex RT-PCR.

The Panther Fusion system compares the fluorescence signal to a predetermined cut-off to produce a qualitative result for the presence or absence of the analyte.

The analytes and the channel used for their detection on the Panther Fusion system is summarized in the table below.

Analyte	Gene Targeted	Instrument Channel
Influenza A Virus	Matrix	FAM
Respiratory Syncytial Virus A/B	Matrix	HEX
SARS-CoV-2	ORF1ab	ROX
Influenza B Virus	Matrix	RED647
Internal Control	Not applicable	RED677

## Warnings and Precautions




- A. For *in vitro* diagnostic use. Carefully read this entire package insert and the *Panther/Panther Fusion System Operator's Manual*.
- B. For professional use.
- C. The Panther Fusion Enhancer Reagent-S (FER-S) is corrosive, harmful if swallowed and causes severe skin burns and eye damage.
- D. Only personnel adequately trained on the use of this assay and in handling potentially infectious materials should perform these procedures. If a spill occurs, immediately disinfect using appropriate site procedures.
- E. Handle all specimens as if infectious using safe laboratory procedures. Refer to Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with 2019-nCoV. <https://www.cdc.gov/coronavirus/2019-ncov/lab/lab-biosafety-guidelines.html>.
- F. Specimens may be infectious. Use Universal Precautions when performing this assay. Proper handling and disposal methods should be established by the laboratory director. Only personnel adequately trained in handling infectious materials should be permitted to perform this diagnostic procedure.<sup>7</sup>

**Note:** *If infection with a novel influenza A virus is suspected based on current clinical and epidemiological screening criteria recommended by public health authorities, collect specimens with appropriate infection control precautions for novel virulent influenza viruses and send to state or local health department for testing. Do not attempt viral culture in these cases unless a BSL 3+ facility is available to receive and culture specimens.*
- G. If infection with 2019-nCoV is suspected based on current clinical screening criteria recommended by public health authorities, specimens should be collected with appropriate infection control precautions.

- H. Use appropriate personal protective equipment when collecting and handling specimens from individuals suspected of being infected with SARS-CoV-2 as outlined in the CDC Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with 2019 Novel Coronavirus (2019-nCoV).
- I. Use only supplied or specified disposable laboratory ware.
- J. Wear disposable, powderless gloves, protective eye wear, and laboratory coats when handling specimens and reagents. Wash hands thoroughly after handling specimens and reagents. Dispose of all material that has come into contact with specimens and reagents in accordance with applicable national, international, and regional regulations.
- K. Expiration dates listed on the Panther Fusion Specimen Lysis Tubes pertain to the transfer of sample into the tube and not to testing of the sample. Specimens collected/transferred any time prior to these expiration dates are valid for testing provided they are transported and stored in accordance with the appropriate package insert, even if these expiration dates have passed.
- L. Maintain proper storage conditions during specimen shipping to ensure the integrity of the specimen. Specimen stability under shipping conditions other than those recommended has not been evaluated.
- M. Avoid cross-contamination during the specimen handling steps. Specimens can contain extremely high levels of virus or other organisms. Ensure that specimen containers do not come in contact with one another, and discard used materials without passing them over any open containers. Change gloves if they come in contact with specimens.
- N. Do not use the reagents and controls after the expiration date.
- O. Store assay components at the recommended storage condition. See *Reagent Storage and Handling Requirements* (page 7), and *Panther Fusion System Test Procedure* (page 12) for more information.
- P. Do not combine any assay reagents or fluids. Do not top off reagents or fluids; the Panther Fusion system verifies reagent levels.
- Q. Avoid microbial and ribonuclease contamination of reagents.
- R. Quality control requirements must be performed in conformance with local, state, and/or federal regulations or accreditation requirements and your laboratory's standard quality control procedures.
- S. Do not use the assay cartridge if the storage pouch has lost its seal or if the assay cartridge foil is not intact. Contact Hologic if either occurs.
- T. Do not use the fluid packs if the foil seal is leaking. Contact Hologic if this occurs.
- U. Handle the assay cartridges with care. Do not drop or invert assay cartridges. Avoid prolonged exposure to ambient light.
- V. Do not use material that may contain Guanidinium thiocyanate or any guanidine-containing materials on the instrument. Highly reactive and/or toxic compounds may form if combined with sodium hypochlorite.

W. Some reagents in the kit are labeled with hazard information.

**Note:** For information on any hazard and precautionary statements that may be associated with reagents, refer to the Safety Data Sheet Library at [www.hologicds.com](http://www.hologicds.com).

	<b>Panther Fusion Oil</b> <i>Polydimethylsiloxane 95-100%</i>  <b>WARNING</b> H315 - Causes skin irritation H319 - Causes serious eye irritation P264 - Wash face, hands and any exposed skin thoroughly after handling P280 - Wear protective gloves/protective clothing/eye protection/face protection
 	<b>Panther Fusion Enhancer Reagent-S</b> <i>Lithium Hydroxide, Monohydrate 5-10%</i>  <b>DANGER</b> H302 - Harmful if swallowed H314 - Causes severe skin burns and eye damage P260 - Do not breathe dust/fume/gas/mist/vapors/spray P264 - Wash face, hands and any exposed skin thoroughly after handling P270 - Do not eat, drink or smoke when using this product P280 - Wear protective gloves/protective clothing/eye protection/face protection

## Reagent Storage and Handling Requirements

A. The following table provides storage and handling requirements for this assay.

Reagent	Unopened Storage	On Board/ Open Stability <sup>1</sup>	Opened Storage
Panther Fusion SARS-CoV-2/Flu A/B/RSV Assay Cartridge	2°C to 8°C	60 days	2°C to 8°C <sup>2</sup>
Panther Fusion Capture Reagent-S (FCR-S)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Enhancer Reagent-S (FER-S)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Internal Control-S (IC-S)	2°C to 8°C	(In wFCR-S)	Not applicable
Panther Fusion Elution Buffer	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Oil	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Reconstitution Buffer I	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion SARS-CoV-2/Flu A/B/RSV Positive Control	2°C to 8°C	Single use vial	Not applicable-single use
Panther Fusion Negative Control	2°C to 8°C	Single use vial	Not applicable-single use

When reagents are removed from the Panther Fusion system, return them immediately to their appropriate storage temperatures.

<sup>1</sup> On board stability starts at the time the reagent is placed on the Panther Fusion system for the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay cartridge, FCR-S, FER-S and IC-S. On board stability starts for the Panther Fusion Reconstitution Buffer I, Panther Fusion Elution Buffer, and Panther Fusion Oil when the reagent pack is first used.

<sup>2</sup> If removed from the Panther Fusion system, store the assay cartridge in an air-tight container with desiccant at the recommended storage temperature.

- B. Working Panther Fusion Capture Reagent-S and Panther Fusion Enhancer Reagent-S are stable for 60 days when capped and stored at 15°C to 30°C. Do not refrigerate.
- C. Discard any unused reagents that have surpassed their on board stability.
- D. Controls are stable until the date indicated on the vials.
- E. Avoid cross-contamination during reagent handling and storage.
- F. **Do not freeze reagents.**

## Specimen Collection and Storage

**Specimens** - Clinical material collected from patient placed in an appropriate transport system. For the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay, this includes NP swab specimens in viral transport medium (VTM), or universal transport medium (UTM).

**Samples** - Represents a more generic term to describe any material for testing on the Panther Fusion system including specimens, specimens transferred into a Panther Fusion Specimen Lysis Tube and controls.

**Note:** Handle all specimens as if they contain potentially infectious agents. Use Universal Precautions.

**Note:** Take care to avoid cross-contamination during specimen handling steps. For example, discard used material without passing over open tubes.

### Specimen Collection

Collect NP swab specimens according to standard technique using a polyester-, rayon, or nylon-tipped swab. Immediately place the swab specimen into 3mL of VTM or UTM.

**Note:** Do not use medium that may contain Guanidium thiocyanate or any guanidine-containing material.

### Specimen Processing

#### *Specimen Processing with the Panther Fusion Specimen Lysis Tube*

1. Prior to testing on the Panther Fusion system, transfer specimen\* collected in UTM or VTM into a Panther Fusion Specimen Lysis Tube.
  - Transfer 500 µL of the specimen to a Panther Fusion Specimen Lysis Tube.

**\*Note:** When testing frozen specimen, allow specimen to reach room temperature prior to processing.

### Specimen Storage

#### *Storing Specimens with the Panther Fusion Specimen Lysis Tube*

1. After collection, specimens can be stored at 2°C to 8°C up to 96 hours before transferred to the Panther Fusion Specimen Lysis Tube. Remaining specimen volumes can be stored at ≤-70°C.
2. Samples (in the Panther Fusion Specimen Lysis Tube) can be stored under the following conditions:
  - 15°C to 30°C up to 6 days or
  - 2°C to 8°C, -20°C, and -70°C for up to 30 days
3. Previously tested samples should be covered with a new, clean plastic film or foil barrier.
4. If assayed samples need to be frozen or shipped, remove the penetrable cap and place a new non-penetrable cap on the specimen tubes. If samples need to be shipped for testing at another facility, recommended temperatures must be maintained. Prior to uncapping previously tested and recapped samples, specimen transport tubes must be centrifuged for 5 minutes at 420 Relative Centrifugal Force (RCF) to bring all of the liquid down to the bottom of the tube. Avoid splashing and cross-contamination.



**Specimen Transport**

Maintain specimen storage conditions as described in the *Specimen Collection and Storage* section on page 8.

**Note:** *Specimens must be shipped in accordance with applicable national, international, and regional transportation regulations.*

## Panther Fusion System

The Panther Fusion System is an integrated nucleic acid testing system that fully automates all steps necessary to perform various Panther Fusion assays from sample processing through amplification, detection, and data reduction.

### Reagents and Materials Provided for Panther Fusion SARS-CoV-2/Flu A/B/RSV Assay

#### Assay Packaging

Components <sup>1</sup>	Part No.	Storage
<b>Panther Fusion SARS-CoV-2/Flu A/B/RSV Assay Cartridges 96 Tests</b> Panther Fusion SARS-CoV-2/Flu A/B/RSV assay cartridge, 12 tests, 8 per box	PRD-07400	2°C to 8°C
<b>Panther Fusion Internal Control-S 960 Tests</b> Panther Fusion Internal Control-S tube, 4 per box	PRD-04332	2°C to 8°C
<b>Panther Fusion SARS-CoV-2/Flu A/B/RSV Controls</b> Panther Fusion SARS-CoV-2/Flu A/B/RSV Positive Controls tube, 5 per box Panther Fusion Negative Control tube, 5 per box	PRD-07401	2°C to 8°C
<b>Panther Fusion Extraction Reagent-S 960 Tests</b> Panther Fusion Capture Reagent-S bottle, 240 tests, 4 per box Panther Fusion Enhancer Reagent-S bottle, 240 tests, 4 per box	PRD-04331	15°C to 30°C
<b>Panther Fusion Elution Buffer 2400 Tests</b> Panther Fusion Elution Buffer pack, 1200 tests, 2 per box	PRD-04334	15°C to 30°C
<b>Panther Fusion Reconstitution Buffer I 1920 Tests</b> Panther Fusion Reconstitution Buffer I pack, 960 tests, 2 per box	PRD-04333	15°C to 30°C
<b>Panther Fusion Oil 1920 Tests</b> Panther Fusion Oil pack, 960 tests, 2 per box	PRD-04335	15°C to 30°C

<sup>1</sup> Components can also be ordered in the following bundles:

Panther Fusion Universal Fluids Kit, PRD-04430, contains 1 each Panther Fusion Oil and Panther Fusion Elution buffer.

Panther Fusion Assay Fluids I-S, PRD-04431, contains 2 Panther Fusion Extraction Reagents-S, 2 Panther Fusion Internal Control-S, and 1 Panther Fusion Reconstitution Buffer I.

#### Individually Packaged Items

Items	Part No.
Panther Fusion Specimen Lysis Tubes, 100 per bag	PRD-04339

**Materials Required and Available Separately**

**Note:** Materials available from Hologic have catalog numbers listed, unless otherwise specified.

<b>Material</b>	<b>Cat. No.</b>
Panther™ System	303095
Panther Fusion System	PRD-04172
Panther Fusion Module	PRD-04173
Aptima™ Assay Fluids Kit (Aptima Wash Solution, Aptima Buffer for Deactivation Fluid, and Aptima Oil Reagent)	303014 (1000 tests)
Multi-tube units (MTUs)	104772-02
Panther Waste Bag Kit	902731
Panther Waste Bin Cover	504405
Or Panther System Run Kit for Real Time Assays contains MTUs, waste bags, waste bin covers, and assay fluids	PRD-03455 (5000 tests)
Or Panther System Run Kit (when running TMA assays in parallel with real time-TMA assays) contains MTUs, waste bags, waste bin covers, auto detect*, and assay fluids	303096 (5000 tests)
Panther Fusion Tube Trays, 1008 tests, 18 trays per box	PRD-04000
Tips, µL, filtered, liquid-sensing, conductive, and disposable.	901121 (10612513 Tecan) 903031 (10612513 Tecan)
Not all products are available in all regions. Contact your representative for region-specific information.	MME-04134 (30180117 Tecan) MME-04128
Aptima penetrable caps (optional)	105668
Replacement non-penetrable caps (optional)	103036A
Replacement extraction reagent bottle caps	CL0040
P1000 pipettor and tips with hydrophobic plugs	-
Bleach, 5% to 8.25% (0.7 M to 1.16 M) sodium hypochlorite solution <b>Note:</b> Refer to the <i>Panther/Panther Fusion System Operator's Manual</i> for instructions on preparing diluted sodium hypochlorite solution.	-
Disposable powderless gloves	-

\*Needed only for Panther Aptima TMA assays.

## Panther Fusion System Test Procedure

**Note:** Refer to the Panther/Panther Fusion System Operator's Manual for additional procedural information.

### A. Work Area Preparation

1. Wipe down work surfaces with 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution. Allow the sodium hypochlorite solution to contact surfaces for at least 1 minute and follow with a deionized (DI) water rinse. Do not allow the sodium hypochlorite solution to dry. Cover the bench surface with clean, plastic-backed absorbent laboratory bench covers.
2. Clean a separate work surface where samples will be prepared using the procedure described in step A.1.

### B. Reagent Preparation

1. Remove the bottles of IC-S, FCR-S and FER-S from storage.
2. Open the bottles of IC-S, FCR-S and FER-S, and discard the caps. Open the TCR door on the upper bay of the Panther Fusion system.
3. Place the IC-S, FCR-S and FER-S bottles in the appropriate positions on the TCR carousel.
4. Close the TCR door.

**Note:** The Panther Fusion system adds the IC-S to the FCR-S. After the IC-S is added to the FCR-S, it is referred to as wFCR-S (working FCR-S). If the FCR-S and FER-S are removed from the system, use new caps and immediately store according to the proper storage conditions.

### C. Specimen Handling

**Note:** Prepare specimens per the Specimen Processing instructions in the Specimen Collection and Storage section before loading specimens onto the Panther Fusion system.

1. **Do not vortex samples.**
2. Inspect sample tubes before loading into the rack. If a sample tube contains bubbles or has a lower volume than is typically observed, gently tap the bottom of the tube to bring contents to the bottom.

**Note:** To avoid a processing error, ensure adequate specimen volume is added to the Panther Fusion Specimen Lysis Tube. When 500  $\mu$ L of NP swab specimen is added to the Panther Fusion Specimen Lysis Tube, there is sufficient volume to perform 3 nucleic acid extractions.

### D. System Preparation

For instructions on setting up the Panther Fusion system including loading samples, reagents, assay cartridges and universal fluids, refer to the Panther/Panther Fusion System Operator's Manual.

**Procedural Notes**

## A. Controls

1. The Panther Fusion SARS-CoV-2/Flu A/B/RSV Positive Control and Panther Fusion Negative Control can be loaded in any rack position, in any Sample Bay lane on the Panther Fusion system.
2. Once the control tubes are pipetted and are processed for the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay, they are active for up to 30 days (control frequency configured by an administrator) unless control results are invalid or a new assay cartridge lot is loaded.
3. Each control tube can be tested once.
4. Patient specimen pipetting begins when one of the following two conditions is met:
  - a. Valid results for the controls are registered on the system.
  - b. A pair of controls is currently in process on the system.

## Quality Control

A run or specimen result may be invalidated by the Panther Fusion system if problems occur while performing the assay. Specimens with invalid results must be retested.

### Negative and Positive Controls

To generate valid results, a set of assay controls must be tested. One replicate of the negative assay control and positive assay control must be tested each time a new lot of assay cartridges is loaded on the Panther Fusion system or when the current set of valid controls for an active cartridge lot have expired.

The Panther Fusion system is configured to require assay controls run at an administrator-specified interval of up to 30 days. Software on the Panther Fusion system alerts the operator when assay controls are required and does not start new tests until the assay controls are loaded and have started processing.

During processing, criteria for acceptance of the assay controls are automatically verified by the Panther Fusion system. To generate valid results, the assay controls must pass a series of validity checks performed by the Panther Fusion system.

If the assay controls pass all validity checks, they are considered valid for the administrator-specified time interval. When the time interval has passed, the assay controls are expired by the Panther Fusion system and requires a new set of assay controls be tested prior to starting any new samples.

If any one of the assay controls fails the validity checks, the Panther Fusion system automatically invalidates the affected samples and requires a new set of assay controls be tested prior to starting any new samples.

### Internal Control

An internal control is added to each sample during the extraction process. During processing, the internal control acceptance criteria is automatically verified by the Panther Fusion system software. Detection of the internal control is not required for samples that are positive for SARS-CoV-2, Flu A, Flu B and/or RSV. The internal control must be detected in all samples that are negative for SARS-CoV-2, Flu A, Flu B, and RSV; samples that fail to meet that criteria will be reported as Invalid. Each sample with an Invalid result must be retested.

The Panther Fusion system is designed to accurately verify processes when procedures are performed following the instructions provided in this package insert and the *Panther/Panther Fusion System Operator's Manual*.

### Interpretation of Results

The Panther Fusion system automatically determines the test results for samples and controls. Results for SARS-CoV-2, Flu A, Flu B, and RSV detection are reported separately. A test result may be negative, positive, or invalid.

Table 1 shows the possible results reported in a valid run with result interpretations.

Table 1: Result Interpretation

SARS-CoV-2 Result	Flu A Result	Flu B Result	RSV Result	IC Result	Interpretation
Neg	Neg	Neg	Neg	Valid	SARS-CoV-2, Flu A, Flu B, and RSV not detected.
Neg	POS	Neg	Neg	Valid	Flu A detected. SARS-CoV-2, Flu B, and RSV not detected.
Neg	Neg	POS	Neg	Valid	Flu B detected. SARS-CoV-2, Flu A, and RSV not detected.
Neg	Neg	Neg	POS	Valid	RSV detected. SARS-CoV-2, Flu A, and Flu B not detected.
POS	Neg	Neg	Neg	Valid	SARS-CoV-2 detected. Flu A, Flu B, and RSV not detected.
Neg	POS	POS	Neg	Valid	Flu A and Flu B detected. SARS-CoV-2 and RSV not detected.
Neg	Neg	POS	POS	Valid	Flu B and RSV detected. SARS-CoV-2 and Flu A not detected.
Neg	POS	Neg	POS	Valid	Flu A and RSV detected. SARS-CoV-2 and Flu B not detected.
POS	POS	Neg	Neg	Valid	SARS-CoV-2 and Flu A detected. Flu B and RSV not detected.
POS	Neg	POS	Neg	Valid	SARS-CoV-2 and Flu B detected. Flu A and RSV not detected.
POS	Neg	Neg	POS	Valid	SARS-CoV-2 and RSV detected. Flu A and Flu B not detected.
Neg	POS	POS	POS	Valid	Flu A, Flu B, and RSV detected. SARS-CoV-2 not detected. Triple infections are rare. Retest to confirm result.
POS	Neg	POS	POS	Valid	SARS-CoV-2, Flu B, and RSV detected. Flu A not detected. Triple infections are rare. Retest to confirm result.
POS	POS	Neg	POS	Valid	SARS-CoV-2, Flu A, and RSV detected. Flu B not detected. Triple infections are rare. Retest to confirm result.
POS	POS	POS	Neg	Valid	SARS-CoV-2, Flu A, and Flu B detected. RSV not detected. Triple infections are rare. Retest to confirm result.
POS	POS	POS	POS	Valid	SARS-CoV-2, Flu A, Flu B, and RSV detected. Quadruple infections are rare. Retest to confirm result.
Invalid	Invalid	Invalid	Invalid	Invalid	Invalid. There was an error in the generation of the result; retest sample.

Note: POS result will be accompanied by cycle threshold (Ct) values.

Note: Detection of internal control is not required for samples that are positive for SARS-CoV-2, Flu A, Flu B, and/or RSV.

**Limitations**

- A. Use of this assay is limited to personnel who are trained in the procedure. Failure to follow these instructions may result in erroneous results.
- B. Reliable results are dependent on adequate specimen collection, transport, storage, and processing.
- C. Avoid contamination by adhering to good laboratory practices and to the procedures specified in this package insert.
- D. Negative results do not preclude SARS-CoV-2, influenza A virus, influenza B virus, or RSV infections and should not be used as the sole basis for treatment or other management decisions.
- E. This test does not differentiate influenza A subtypes (i.e. H1N1, H3N2) or RSV subgroups (i.e., A or B); additional testing is required to differentiate any specific influenza A subtypes or strains or specific RSV subgroups, in consultation with local public health departments.
- F. A positive result indicates the detection of nucleic acid from the relevant virus. Nucleic acid may persist even after the virus is no longer viable.



## SARS-CoV-2/Flu A/B/RSV Assay Performance

### Analytical Sensitivity

The analytical sensitivity (limit of detection or LoD) of the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was determined by testing dilutions of pooled negative clinical nasopharyngeal (NP) swab VTM/UTM matrix spiked with either the WHO International Standard for SARS-CoV-2, NIBSC (20/146) or the following virus cultures: Influenza A (2 strains), Influenza B (2 strains), RSV A and RSV B (1 strain each). A minimum of 24 replicates were tested with each of three reagent lots for a combined total of a minimum of 72 replicates per dilution. Each target specific LoD concentration was confirmed by testing an additional 24 replicates in negative clinical NP swab VTM/UTM matrix with one reagent lot. The LoD for each was determined by Probit analysis and the highest value between three reagent lots is summarized in Table 2.

Table 2: Analytical Sensitivity

Viral Strain/Standard	LoD Concentration
WHO International Standard SARS-CoV-2, NIBSC (20/146)	47.20 IU/mL
SARS-CoV-2 USA-WA1/2020	0.03 TCID <sub>50</sub> /mL
Influenza A/Brisbane/02/18 (H1N1)	0.06 TCID <sub>50</sub> /mL
Influenza A/Kansas/14/17 (H3N2)	0.11 TCID <sub>50</sub> /mL
Influenza B/Washington/02/19 (Victoria lineage)	0.03 TCID <sub>50</sub> /mL
Influenza B/Phuket/3073/13 (Yamagata lineage)	0.002 TCID <sub>50</sub> /mL
RSV A	0.02 TCID <sub>50</sub> /mL
RSV B	0.03 TCID <sub>50</sub> /mL

## Reactivity

The reactivity of the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was determined by testing virus strains in negative clinical NP swab VTM/UTM matrix. Each strain was tested in triplicate with one reagent lot. Table 3 shows the lowest concentration of each strain in which 100% positivity was observed.

Table 3: Analytical Reactivity Summary for SARS-CoV-2, Flu A and Flu B and RSV Strains

Description	Subtype	Concentration	SARS-CoV-2	Flu A	Flu B	RSV
USA-WA1/2020*	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA-CA1/2020	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA-AZ1/2020	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA-WI1/2020	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/OR-OHSU-PHL00037/ 2021 B.1.1.7	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
Uganda/MUWRP-20200195568/ 2020 A.23.1	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/PHC658/2021 B.1.617.2	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/MD-HP05285/2021 B.1.617.2	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/CA/VRLC009/2021 B.1.427	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/CA/VRLC012/2021 P.2	SARS-CoV-2	0.3 TCID <sub>50</sub> /mL	+	-	-	-
USA/MD-HP03056/2021 B.1.525	SARS-CoV-2	0.3 TCID <sub>50</sub> /mL	+	-	-	-
USA/CA-Stanford-16_S02/ 2021 B.1.617.1	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
Peru/un-CDC-2-4069945/ 2021 C.37	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/MD-HP20874/2021 B.1.1.529	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
USA/GA-EHC-2811C/ 2021 B.1.1.529	SARS-CoV-2	0.09 TCID <sub>50</sub> /mL	+	-	-	-
A/Brisbane/02/18*	Flu A (H1N1)	0.18 TCID <sub>50</sub> /mL	-	+	-	-
A/Michigan/45/2015	Flu A (H1N1)	0.18 TCID <sub>50</sub> /mL	-	+	-	-
A/Christ Church/16/2010	Flu A (H1N1)	180 TCID <sub>50</sub> /mL	-	+	-	-
A/Kentucky/2/06	Flu A (H1N1)	1.8 TCID <sub>50</sub> /mL	-	+	-	-
A/Solomon Islands/03/06	Flu A (H1N1)	1.8 TCID <sub>50</sub> /mL	-	+	-	-
A/Guangdong-maonan/1536/2019	Flu A (H1N1)	180 TCID <sub>50</sub> /mL	-	+	-	-
A/Taiwan/42/2006	Flu A (H1N1)	1.8 TCID <sub>50</sub> /mL	-	+	-	-
A/Henan/8/05	Flu A (H1N1)	1.8 TCID <sub>50</sub> /mL	-	+	-	-
A/Hawaii/15/01	Flu A (H1N1)	18 TCID <sub>50</sub> /mL	-	+	-	-
A/California/07/2009	Flu A (H1N1)	0.18 TCID <sub>50</sub> /mL	-	+	-	-

Table 3: Analytical Reactivity Summary for SARS-CoV-2, Flu A and Flu B and RSV Strains (Continued)

Description	Subtype	Concentration	SARS-CoV-2	Flu A	Flu B	RSV
A/Hawaii/66/2019	Flu A (H1N1)	180 CEID <sub>50</sub> /mL	-	+	-	-
A/Indiana/02/2020	Flu A (H1N1)	60 CEID <sub>50</sub> /mL	-	+	-	-
A/Michigan/45/2015	Flu A (H1N1)	1.8 TCID <sub>50</sub> /mL	-	+	-	-
A/Kansas/14/17*	Flu A (H3N2)	0.33 TCID <sub>50</sub> /mL	-	+	-	-
A/Arizona/45/2018	Flu A (H3N2)	3.3 FFU/mL	-	+	-	-
A/New York/21/2020	Flu A (H3N2)	3.3 FFU/mL	-	+	-	-
A/Hong Kong/45/2019	Flu A (H3N2)	3.3 FFU/mL	-	+	-	-
A/Singapore/INFIMH-16-0019/2016	Flu A (H3N2)	110 CEID <sub>50</sub> /mL	-	+	-	-
A/Hong Kong/2671/2019	Flu A (H3N2)	33 TCID <sub>50</sub> /mL	-	+	-	-
A/Hiroshima/52/05	Flu A (H3N2)	3.3 TCID <sub>50</sub> /mL	-	+	-	-
A/Costa Rica/07/99	Flu A (H3N2)	33 TCID <sub>50</sub> /mL	-	+	-	-
A/Port Chalmers/1/73	Flu A (H3N2)	3.3 TCID <sub>50</sub> /mL	-	+	-	-
A/Brazil/113/99	Flu A (H3N2)	3.3 TCID <sub>50</sub> /mL	-	+	-	-
A/Perth/16/2009	Flu A (H3N2)	0.33 TCID <sub>50</sub> /mL	-	+	-	-
A/Texas/50/2012	Flu A (H3N2)	0.33 TCID <sub>50</sub> /mL	-	+	-	-
A/Hong Kong/4801/2014	Flu A (H3N2)	3.3 TCID <sub>50</sub> /mL	-	+	-	-
A/Indiana/08/2011	Flu A (H3N2)	3.3 TCID <sub>50</sub> /mL	-	+	-	-
A/Hong Kong/486/97	Flu A (H5N1)	0.01 ng/mL	-	+	-	-
B/Washington/02/2019*	Flu B (Victoria)	0.09 TCID <sub>50</sub> /mL	-	-	+	-
B/Colorado/06/2017	Flu B (Victoria)	0.09 TCID <sub>50</sub> /mL	-	-	+	-
B/Florida/78/2015	Flu B (Victoria)	0.9 TCID <sub>50</sub> /mL	-	-	+	-
B/Alabama/2/17	Flu B (Victoria)	0.09 TCID <sub>50</sub> /mL	-	-	+	-
B/Ohio/1/2005	Flu B (Victoria)	0.9 TCID <sub>50</sub> /mL	-	-	+	-
B/Michigan/09/2011	Flu B (Victoria)	3 TCID <sub>50</sub> /mL	-	-	+	-
B/Hawaii/01/2018 (NA D197N)	Flu B (Victoria)	0.9 TCID <sub>50</sub> /mL	-	-	+	-
B/Brisbane/33/08	Flu B (Victoria)	0.09 TCID <sub>50</sub> /mL	-	-	+	-
B/Phuket/3073/2013*	Flu B (Yamagata)	0.006 TCID <sub>50</sub> /mL	-	-	+	-
B/Wisconsin/1/2010	Flu B (Yamagata)	2 TCID <sub>50</sub> /mL	-	-	+	-
B/Utah/9/14	Flu B (Yamagata)	0.006 TCID <sub>50</sub> /mL	-	-	+	-
B/St. Petersburg/04/06	Flu B (Yamagata)	0.06 TCID <sub>50</sub> /mL	-	-	+	-
B/Texas/81/2016	Flu B (Yamagata)	20 TCID <sub>50</sub> /mL	-	-	+	-
B/Indiana/17/2017	Flu B (Yamagata)	0.6 TCID <sub>50</sub> /mL	-	-	+	-

Table 3: Analytical Reactivity Summary for SARS-CoV-2, Flu A and Flu B and RSV Strains (Continued)

Description	Subtype	Concentration	SARS-CoV-2	Flu A	Flu B	RSV
B/Oklahoma/10/2018	Flu B (Yamagata)	2 TCID <sub>50</sub> /mL	-	-	+	-
B/Massachusetts/02/2012	Flu B (Yamagata)	0.2 TCID <sub>50</sub> /mL	-	-	+	-
B/Lee/40	Flu B	0.09 TCID <sub>50</sub> /mL	-	-	+	-
RSV-A/2006 Isolate*	RSVA	0.06 TCID <sub>50</sub> /mL	-	-	-	+
RSV A/4/2015 isolate #1	RSVA	0.06 TCID <sub>50</sub> /mL	-	-	-	+
RSV A/A2	RSVA	0.06 TCID <sub>50</sub> /mL	-	-	-	+
RSV A/12/2014 isolate #2	RSVA	0.06 TCID <sub>50</sub> /mL	-	-	-	+
RSV-B/CH93(18)-18*	RSVB	0.3 TCID <sub>50</sub> /mL	-	-	-	+
RSV B/3/2015 isolate #1	RSVB	0.09 TCID <sub>50</sub> /mL	-	-	-	+
RSV B/9320	RSVB	0.09 TCID <sub>50</sub> /mL	-	-	-	+

\*Strain used to establish LoD.

### Analytical Specificity and Microbial Interference

Analytical specificity (cross-reactivity) and microbial interference with the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was evaluated in the presence of closely related and non-targeted organisms. Panels consisting of 41 organisms (Table 4) were tested in pooled negative clinical NP swab VTM/UTM matrix in the absence or presence of 3x LoD SARS-CoV-2, Flu A, Flu B and RSV. Bacteria were tested at 10<sup>6</sup> CFU/mL and viruses were tested at 10<sup>5</sup> TCID<sub>50</sub>/mL, except where noted. No cross-reactivity or microbial interference was observed for any of the 41 organisms tested in the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay at the following concentrations.

Table 4: Cross Reactivity and Microbial Interference Microorganisms

Microorganism	Concentration <sup>1</sup>	Microorganism	Concentration <sup>1</sup>
Adenovirus 1	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Bordetella pertussis</i>	1x10 <sup>6</sup> CFU/mL
Adenovirus 7a	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Candida albicans</i>	1x10 <sup>6</sup> CFU/mL
CMV Strain AD 169	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Chlamydomphila pneumoniae</i>	1x10 <sup>6</sup> IFU/mL
Human coronavirus 229E	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Corynebacterium diphtheriae</i>	1x10 <sup>6</sup> CFU/mL
Human coronavirus NL63	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Escherichia coli</i>	1x10 <sup>6</sup> CFU/mL
Human coronavirus OC43	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Haemophilus influenzae</i>	1x10 <sup>6</sup> CFU/mL
Epstein-Barr virus (EBV)	1x10 <sup>6</sup> copies/mL	<i>Lactobacillus plantarum</i>	1x10 <sup>6</sup> CFU/mL
Enterovirus (e.g. EV68)	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Legionella pneumophila</i>	1x10 <sup>6</sup> CFU/mL
Human coronavirus HKU1 <sup>2</sup>	1x10 <sup>6</sup> copies/mL	<i>Moraxella catarrhalis</i>	1x10 <sup>5</sup> CFU/mL
Human Metapneumovirus (hMPV)	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Mycobacterium tuberculosis</i>	1x10 <sup>9</sup> rRNA copies/mL
HPIV-1	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Mycoplasma pneumoniae</i>	1x10 <sup>9</sup> rRNA copies/mL
HPIV-2	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Neisseria spp</i>	1x10 <sup>6</sup> CFU/mL
HPIV-3	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Neisseria meningitides</i>	1x10 <sup>6</sup> CFU/mL
HPIV-4	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Neisseria mucosa</i>	1x10 <sup>6</sup> CFU/mL
Measles	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Pneumocystis jirovecii</i>	1x10 <sup>6</sup> CFU/mL
MERS-Coronavirus	5x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Pseudomonas aeruginosa</i>	1x10 <sup>6</sup> CFU/mL
Mumps virus	1x10 <sup>5</sup> TCID <sub>50</sub> /mL	<i>Staphylococcus aureus</i>	1x10 <sup>6</sup> CFU/mL
Rhinovirus 1A	1x10 <sup>4</sup> TCID <sub>50</sub> /mL	<i>Staphylococcus epidermidis</i>	1x10 <sup>6</sup> CFU/mL
SARS coronavirus 1 <sup>2</sup>	1x10 <sup>6</sup> copies/mL	<i>Streptococcus pneumoniae</i>	1x10 <sup>6</sup> CFU/mL
Varicella Zoster Virus	1x10 <sup>3</sup> TCID <sub>50</sub> /mL	<i>Streptococcus pyogenes</i>	1x10 <sup>6</sup> CFU/mL
		<i>Streptococcus salivarius</i>	1x10 <sup>6</sup> CFU/mL

<sup>1</sup>CFU = Colony Forming Units; IFU = Inclusion Forming Units; TCID<sub>50</sub> = Median Tissue Culture Infectious Dose

<sup>2</sup>Cultured virus and whole genome purified nucleic acid for Human HKU1 and SARS-coronavirus are not readily available. HKU1 and SARS-coronavirus *in vitro* transcript (IVT) corresponding to the ORF1a gene regions targeted by the assay were used to evaluate cross-reactivity and microbial interference.

## Competitive Interference

Competitive interference in the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was evaluated using pairs of targeted viruses at low/high concentrations in pooled negative clinical NP swab VTM/UTM matrix. The low concentration was tested at 3x LoD, while the high concentration virus was tested at 1000x LoD. Results of the study are shown in Table 5. The presence of two viruses at varying concentrations had no effect on the analytical sensitivity of one target in the presence of high concentrations of the other target.

Table 5: Competitive Interference

Low Target		High Target		SARS-CoV-2 (detected)	Flu A (detected)	Flu B (detected)	RSV (detected)
Virus	3x LoD (TCID <sub>50</sub> /mL)	Virus	1000x LoD (TCID <sub>50</sub> /mL)				
SARS-CoV-2	0.09	Flu A	110	+	+	-	-
SARS-CoV-2	0.09	Flu B	30	+	-	+	-
SARS-CoV-2	0.09	RSV	30	+	-	-	+
Flu A	0.33	SARS-CoV-2	30	+	+	-	-
Flu A	0.33	Flu B	30	-	+	+	-
Flu A	0.33	RSV	30	-	+	-	+
Flu B	0.09	SARS-CoV-2	30	+	-	+	-
Flu B	0.09	Flu A	110	-	+	+	-
Flu B	0.09	RSV	30	-	-	+	+
RSV	0.09	SARS-CoV-2	30	+	-	-	+
RSV	0.09	Flu A	110	-	+	-	+
RSV	0.09	Flu B	30	-	-	+	+

## Interference

Interfering endogenous and exogenous substances (mucin, whole blood, other potential medications and over-the-counter products) that may be present in a specimen were evaluated in the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay. Clinically relevant concentrations of potentially interfering substances were added to pooled clinical negative NP swab VTM/UTM matrix and tested in the absence and presence of SARS-CoV-2, Flu A, Flu B and RSV cultured virus at their respective 3x LoD concentrations. The substances and concentrations are shown in Table 6.

No impact on the performance of the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was seen for any of the substances at the concentrations tested.

Table 6: Potentially Interfering Substances

Substance Type	Substance Name	Active Ingredient(s)	Concentration <sup>1</sup>
Endogenous	Mucin	Purified mucin protein	60 µg/mL
	Blood (human)	N/A	2% v/v
Nasal sprays or drops	Neo-Syneprine®	Phenylephrine	15% v/v
	Anefrin	Oxymetazoline	15% v/v
	Saline	Sodium chloride	15% v/v
	Ventolin HFA <sup>2</sup>	Albuterol	45 ng/mL
Nasal corticosteroids	QVAR® Beconase AQ <sup>2</sup>	Beclomethasone	15 ng/mL
	Dexacort <sup>2</sup>	Dexamethasone	12 µg/mL
	Nasacort	Triamcinolone	5% v/v
	Flonase	Fluticasone	5% v/v
	Rhinocort	Budesonide	5% v/v
	Nasonex <sup>2</sup>	Mometasone	0.5 ng/mL
	AEROSPAN® <sup>2</sup>	Flunisolide	10 µg/mL
Nasal gel	Zicam® (Allergy Relief)	Luffa operculata, Galphimia, Glauca, Histaminum hydrochloricum, Sulfur	5% v/v
Throat lozenge	Cepacol Extra Strength	Benzocaine, Menthol	0.7 mg/mL
Anti-viral drug	Relenza® <sup>2</sup>	Zanamivir	3.3 mg/mL
	TamiFlu <sup>2</sup>	Oseltamivir	400 µg/mL
	Virazole <sup>2</sup>	Ribavirin	10.5 µg/mL
Antibiotic, nasal ointment	Bactroban cream <sup>2</sup>	Mupirocin	1.6 µg/mL
Antibiotic, systemic	Tobramycin	Tobramycin	33.1 µg/mL

<sup>1</sup> v/v: volume by volume<sup>2</sup> Active ingredients tested

**Assay Precision**

Panther Fusion SARS-CoV-2/Flu A/B/RSV assay within-lab precision was evaluated with a 5-member panel consisting of virus in negative clinical NP swab VTM/UTM matrix. The panels were tested by two operators on two runs per day, using three reagent lots on three Panther Fusion systems over twelve days.

The panel members are described in Table 7, along with a summary of the agreement with the expected results and the Ct mean and variability analysis between reagent lots, operators, instruments, between and within runs, and overall (total).

*Table 7: Signal Variability of the Panther Fusion SARS-CoV-2/Flu A/B/RSV Assay by Panel Member*

Panel	Description	Analyte	Agreed/N*	Agreement (%)	Mean Ct	Between Lots		Between Instrument		Between Operators		Between Days		Between Runs		Within Run		Total	
						SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)
1	Neg	Internal Control	95/96	99	33.7	0.19	0.57	0.08	0.23	0.00	0.00	0.00	0.00	0.21	0.62	0.29	0.86	0.42	1.23
2	SARS-CoV-2/Flu A Low Pos	Flu A	96/96	100	35.1	0.33	0.93	0.06	0.17	0.00	0.00	0.00	0.00	0.30	0.85	0.56	1.59	0.72	2.04
		SARS-CoV-2	96/96	100	35.9	0.00	0.00	0.13	0.36	0.00	0.00	0.00	0.00	0.00	0.00	0.60	1.67	0.61	1.71
3	Flu B/RSV Low Pos	Flu B	96/96	100	36.0	0.14	0.40	0.09	0.25	0.00	0.00	0.00	0.00	0.00	0.00	0.36	0.99	0.39	1.09
		RSV	96/96	100	36.1	0.12	0.33	0.28	0.77	0.00	0.00	0.00	0.00	0.37	1.04	0.53	1.46	0.71	1.97
4	SARS-CoV-2/Flu A Mod Pos	Flu A	96/96	100	33.9	0.23	0.66	0.00	0.00	0.00	0.00	0.19	0.56	0.00	0.00	0.47	1.37	0.55	1.63
		SARS-CoV-2	96/96	100	34.7	0.21	0.62	0.16	0.45	0.06	0.17	0.00	0.00	0.00	0.00	0.45	1.30	0.52	1.51
5	Flu B/RSV Mod Pos	Flu B	96/96	100	34.7	0.15	0.44	0.00	0.00	0.00	0.00	0.00	0.00	0.06	0.18	0.28	0.80	0.32	0.93
		RSV	96/96	100	34.5	0.10	0.30	0.18	0.51	0.00	0.00	0.00	0.00	0.00	0.00	0.40	1.15	0.44	1.29

\*Agreement to expected panel positivity result.

Low Pos = Low positive 1-2x LoD.

Mod Pos = Moderate positive 3-5x LoD.

Note: Variability from some factors may be numerically negative, which can occur if the variability due to those factors is very small. When this occurs, SD=0 and CV=0%.



## Clinical Performance

The clinical performance of the Panther Fusion SARS-CoV-2/Flu A/B/RSV assay was evaluated in comparison to an FDA Emergency Use Authorization (EUA) nucleic acid amplification test (NAAT) assay and an FDA-cleared Flu/RSV NAAT assay using individual remnant clinical NP specimens in VTM/UTM collected from patients with signs and symptoms of respiratory infection. For the evaluation, a combination of negative, SARS-CoV-2 positive, Flu A positive, Flu B positive, and RSV positive specimens were tested with each assay.

The Positive Percent Agreement (PPA) and Negative Percent Agreement (NPA) for SARS-CoV-2 was calculated in relation to the FDA EUA authorized NAAT assay as the reference result, as shown in Table 8. The assay showed positive and negative percent agreements of 98.1% and 98.5% respectively for SARS-CoV-2.

For Flu A, Flu B and RSV, the PPA and NPA were calculated in relation to the FDA-cleared Flu/RSV NAAT assay as the reference result, as shown in Table 9 for Flu A, Table 10 for Flu B and Table 11 for RSV. The assay showed positive and negative percent agreements of 100.0% and 99.6% respectively, for Flu A, 98.1% and 99.6% for Flu B and 98.1% and 100.0% for RSV.

Table 8: Clinical Performance for SARS-CoV-2

SARS-CoV-2		FDA EUA Authorized NAAT Assay		
		Positive	Negative	Total
Panther Fusion SARS/Flu A/B/RSV Assay	Positive	52	4	56
	Negative	1	256	257
	Total	53	260	313
Positive Agreement (95% CI)		98.1%	(90.1% - 99.7%)	
Negative Agreement (95% CI)		98.5%	(96.1% - 99.4%)	

Table 9: Clinical Performance for Flu A

Flu A		FDA-Cleared Assay		
		Positive	Negative	Total
Panther Fusion SARS/Flu A/B/RSV Assay	Positive	52	1	53
	Negative	0	260	260
	<b>Total</b>	52	261	313
<b>Positive Agreement (95% CI)</b>		100.0%	(93.1% - 100.0%)	
<b>Negative Agreement (95% CI)</b>		99.6%	(97.9% - 99.9%)	

Table 10: Clinical Performance for Flu B

Flu B		FDA-Cleared Assay		
		Positive	Negative	Total
Panther Fusion SARS/Flu A/B/RSV Assay	Positive	52	1	53
	Negative	1	259	260
	<b>Total</b>	53	260	313
<b>Positive Agreement (95% CI)</b>		98.1%	(90.1% - 99.7%)	
<b>Negative Agreement (95% CI)</b>		99.6%	(97.9% - 99.9%)	

Table 11: Clinical Performance for RSV

RSV		FDA-Cleared Assay		
		Positive	Negative	Total
Panther Fusion SARS/Flu A/B/RSV Assay	Positive	52	0	52
	Negative	1	260	261
	<b>Total</b>	53	260	313
<b>Positive Agreement (95% CI)</b>		98.1%	(90.1% - 99.7%)	
<b>Negative Agreement (95% CI)</b>		100.0%	(98.5% - 100.0%)	

## Bibliography

1. Centers for Disease Control and Prevention. <https://www.cdc.gov/2019-ncov/symptoms-testing/symptoms.html>. Accessed August 17, 2021.
2. Centers for Disease Control and Prevention. <https://www.cdc.gov/rsv/about/symptoms.html>. Accessed August 17, 2021.
3. Centers for Disease Control and Prevention. <https://www.cdc.gov/flu/professionals/antivirals/summary-clinicians.htm>.
4. Akers IE, Weber R, Sax H, Böni J, Trkola A, Kuster SP. Influence of time to diagnosis of severe influenza on antibiotic use, length of stay, isolation precautions, and mortality: a retrospective study. *Influenza Other Respir Viruses*. 2017;11(4):337-344. doi:10.1111/irv.12454.
5. Couch, R.B. and Kasel, J.A. 1995. Influenza in *Diagnostic Procedures for Viral, Rickettsial, and Chlamydial Infections*. 7<sup>th</sup> Edition. 431-446.
6. Harper, S.A., Fukuda, K., Uyeki, T.M., Cox, N.J., and Bridges, C.B. 2005. Prevention and Control of Influenza. *MMWR*. 54(RR08):1-40.
7. World Health Organization. Influenza (Seasonal). <http://www.who.int/mediacentre/factsheets/fs211/en/>.
8. Centers for Disease Control and Prevention. Respiratory Syncytial Virus (RSV) Research & Surveillance. <https://www.cdc.gov/rsv/research/us-surveillance.html>. Accessed August 30, 2021.
9. Centers for Disease Control and Prevention. <https://www.cdc.gov/flu/symptoms/index.html>. Accessed August 17, 2021.
10. Centers for Disease Control and Prevention. <https://www.cdc.gov/coronavirus/2019-ncov/symptoms-testing/symptoms.html>. Accessed August 17, 2021.
11. Cucinotta D, Vanelli M. WHO Declares COVID-19 a Pandemic. *Acta Biomed*. 2020 Mar 19;91(1):157-160. doi: 10.23750/abm.v91i1.9397. PMID: 32191675; PMCID: PMC7569573.



**Hologic BV**  
Da Vincilaan 5  
1930 Zaventem  
Belgium

Hologic, Inc.  
10210 Genetic Center Drive  
San Diego, CA 92121 USA

Customer Support: +1 800 442 9892  
customersupport@hologic.com

Technical Support: +1 888 484 4747  
molecularsupport@hologic.com

For more contact information visit [www.hologic.com](http://www.hologic.com).

Hologic, Aptima, Panther, and Panther Fusion are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at [www.hologic.com/patents](http://www.hologic.com/patents).

©2022 Hologic, Inc. All rights reserved.

AW-25328-001 Rev. 001  
2022-05